



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB127446

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/22/2023	13:35
2	CAL2	1	Water	NA	NA	20.5	7.00	09/22/2023	13:36
3	CAL3	1	Water	NA	NA	20.6	10.02	09/22/2023	13:37
4	ICV	1	Water	NA	NA	20.5	7.01	09/22/2023	13:40
5	CCV1	1	Water	NA	NA	20.6	2.01	09/22/2023	15:10
6	O4563-04	1	Solid	20.02	20	20.4	6.15	09/22/2023	15:15
7	O4563-04DUP	1	Solid	20.03	20	20.5	6.16	09/22/2023	15:16
8	O4563-06	1	Solid	20.01	20	20.3	4.15	09/22/2023	15:20
9	CCV2	1	Water	NA	NA	20.6	12.02	09/22/2023	15:25

WORKLIST(Hardcopy Internal Chain)

VB 127446

WorkList Name : CORROSIVITY O4532 WorkList ID : 173880 Department : Wet-Chemistry Date : 09-22-2023 13:27:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4563-04	BEL-23-0007	Solid	Corrosivity	Cool 4 deg C	PSEG03	N11	09/21/2023	9045D
O4563-06	BEL-23-0005	Solid	Corrosivity	Cool 4 deg C	PSEG03		09/21/2023	9045D

Date/Time 09.22.23 13:30
 Raw Sample Received by: JWC
 Raw Sample Relinquished by: CP

Date/Time 09.22.23 17:00
 Raw Sample Received by: CP
 Raw Sample Relinquished by: JWC