


L3127456

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9583e-004

C= -4.4346e-002

Rho= 0.9994964

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/610
0.0	0.000	0.005	0.005	253	0.000	253	0				1
0.2	0.200	0.170	-0.030	1092	0.0 %	1092					16
2.5	2.500	2.585	0.085	13425	0.0 %	13425					3
5.0	5.000	4.840	-0.160	24941	0.0 %	24941					3
7.5	7.500	7.698	0.198	39534	0.0 %	39534					3
10.0	10.000	9.903	-0.097	50795	0.0 %	50795					1

LB127456 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Sep 2023 12:05:07

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	253	-	0.0000	7470A	PPB	25 Sep 2023 12:15:47	S	Std
0.2 - 1	50.2	1092	-	0.2000	7470A	PPB	25 Sep 2023 12:20:09	S	Std
2.5 - 1	52.5	13425	-	2.5000	7470A	PPB	25 Sep 2023 12:22:26	S	Std
5.0 - 1	55	24941	-	5.0000	7470A	PPB	25 Sep 2023 12:24:43	S	Std
7.5 - 1	57.5	39534	-	7.5000	7470A	PPB	25 Sep 2023 12:26:59	S	Std
10.0 - 1	511	50795	-	10.0000	7470A	PPB	25 Sep 2023 12:29:15	S	Std
ICV79 - 1	ICV79	19027	3.6818	-	7470A	PPB	25 Sep 2023 12:32:29	U	SMPL
ICB79 - 1	ICB79	256	0.0058	-	7470A	PPB	25 Sep 2023 12:34:44	U	SMPL
CCV52 - 1	CCV52	23998	4.6552	-	7470A	PPB	25 Sep 2023 12:36:59	U	SMPL
CCB52 - 1	CCB52	246	0.0038	-	7470A	PPB	25 Sep 2023 12:39:14	U	SMPL
CRA - 1	CRA	1069	0.1650	-	7470A	PPB	25 Sep 2023 12:41:28	U	SMPL
HighStd - 1	HighStd	50270	9.8002	-	7470A	PPB	25 Sep 2023 12:43:43	U	SMPL
ChkStd - 1	ChkStd	36990	7.1995	-	7470A	PPB	25 Sep 2023 12:45:58	U	SMPL
PB155776BL - 1	PBW	367	0.0275	-	7470A	PPB	25 Sep 2023 12:51:02	U	SMPL
PB155776BS - 1	LCSW	18469	3.5725	-	7470A	PPB	25 Sep 2023 12:53:18	U	SMPL
O4510-01 - 1	SB-14	519	0.0573	-	7470A	PPB	25 Sep 2023 12:55:33	U	SMPL
O4510-01DUP - 1	SB-14DUP	550	0.0634	-	7470A	PPB	25 Sep 2023 12:57:49	U	SMPL
O4510-01MS - 1	SB-14MS	21299	4.1267	-	7470A	PPB	25 Sep 2023 13:00:06	U	SMPL
O4510-01MSD - 1	SB-14MSD	19524	3.7791	-	7470A	PPB	25 Sep 2023 13:02:23	U	SMPL
O4510-02 - 1	SB-13	1105	0.1720	-	7470A	PPB	25 Sep 2023 13:04:38	U	SMPL
CCV53 - 1	CCV53	24240	4.7026	-	7470A	PPB	25 Sep 2023 13:06:53	U	SMPL
CCB53 - 1	CCB53	212	-0.0028	-	7470A	PPB	25 Sep 2023 13:09:09	U	SMPL
O4521-02 - 1	350	1012	0.1538	-	7470A	PPB	25 Sep 2023 13:11:25	U	SMPL
O4531-02 - 1	TP-BCD	860	0.1241	-	7470A	PPB	25 Sep 2023 13:13:40	U	SMPL
O4531-05 - 1	TP-A	645	0.0820	-	7470A	PPB	25 Sep 2023 13:15:55	U	SMPL
O4537-01 - 1	COMP-1	1894	0.3266	-	7470A	PPB	25 Sep 2023 13:18:11	U	SMPL
O4537-03 - 1	COMP-3	658	0.0845	-	7470A	PPB	25 Sep 2023 13:20:27	U	SMPL
O4537-05 - 1	5873	760	0.1045	-	7470A	PPB	25 Sep 2023 13:22:44	U	SMPL
O4537-06 - 1	5811	842	0.1205	-	7470A	PPB	25 Sep 2023 13:25:00	U	SMPL
O4537-07 - 1	5841	886	0.1292	-	7470A	PPB	25 Sep 2023 13:27:17	U	SMPL
O4563-04 - 1	BEL-23-0007	1024	0.1562	-	7470A	PPB	25 Sep 2023 13:29:33	U	SMPL
O4563-06 - 1	BEL-23-0005	899	0.1317	-	7470A	PPB	25 Sep 2023 13:31:49	U	SMPL
CCV54 - 1	CCV54	23964	4.6486	-	7470A	PPB	25 Sep 2023 13:34:05	U	SMPL
CCB54 - 1	CCB54	226	-0.0001	-	7470A	PPB	25 Sep 2023 13:36:20	U	SMPL
PB155777BL - 1	PBW	522	0.0579	-	7470A	PPB	25 Sep 2023 13:38:36	U	SMPL
PB155777BS - 1	LCSW	17615	3.4052	-	7470A	PPB	25 Sep 2023 13:44:25	U	SMPL
O4537-02 - 1	COMP-2	512	0.0559	-	7470A	PPB	25 Sep 2023 13:46:40	U	SMPL
O4537-04 - 1	5866	409	0.0357	-	7470A	PPB	25 Sep 2023 13:48:56	U	SMPL
O4563-02 - 1	BEL-23-0006	644	0.0818	-	7470A	PPB	25 Sep 2023 13:51:11	U	SMPL
O4563-03 - 1	BEL-23-0008	450	0.0438	-	7470A	PPB	25 Sep 2023 13:53:27	U	SMPL
O4563-03DUP - 1	BEL-23-0008DUP	476	0.0489	-	7470A	PPB	25 Sep 2023 13:55:43	U	SMPL
O4563-03MS - 1	BEL-23-0008MS	16219	3.1319	-	7470A	PPB	25 Sep 2023 13:58:00	U	SMPL
O4563-03MSD - 1	BEL-23-0008MSD	15580	3.0067	-	7470A	PPB	25 Sep 2023 14:00:17	U	SMPL
PB155778BL - 1	PBW	463	0.0463	-	7470A	PPB	25 Sep 2023 14:02:33	U	SMPL
CCV55 - 1	CCV55	23589	4.5752	-	7470A	PPB	25 Sep 2023 14:04:48	U	SMPL
CCB55 - 1	CCB55	147	-0.0156	-	7470A	PPB	25 Sep 2023 14:07:04	U	SMPL
PB155778BS - 1	LCSW	17846	3.4505	-	7470A	PPB	25 Sep 2023 14:12:00	U	SMPL
O4540-01 - 1	CENTRAL-M-R-HEATER	565	0.0663	-	7470A	PPB	25 Sep 2023 14:14:16	U	SMPL
O4540-01DUP - 1	CENTRAL-M-R-HEATERDUP	637	0.0804	-	7470A	PPB	25 Sep 2023 14:16:31	U	SMPL
O4540-01MS - 1	CENTRAL-M-R-HEATERMS	19388	3.7525	-	7470A	PPB	25 Sep 2023 14:18:47	U	SMPL
O4540-01MSD - 1	CENTRAL-M-R-HEATERMSD	17100	3.3044	-	7470A	PPB	25 Sep 2023 14:21:03	U	SMPL
O4540-02 - 1	EMBRIDGE-M-R-HEATER	492	0.0520	-	7470A	PPB	25 Sep 2023 14:23:19	U	SMPL
O4540-03 - 1	TRANSCO-M-R-HEATER	535	0.0604	-	7470A	PPB	25 Sep 2023 14:25:36	U	SMPL
O4566-11 - 1	FB-9-20	297	0.0138	-	7470A	PPB	25 Sep 2023 14:27:52	U	SMPL
O4574-01 - 1	FRAC-TANK-A2421	4212	0.7805	-	7470A	PPB	25 Sep 2023 14:30:12	U	SMPL
O4580-05 - 1	ORANGE-WATER	454	0.0446	-	7470A	PPB	25 Sep 2023 14:32:28	U	SMPL
CCV56 - 1	CCV56	23422	4.5424	-	7470A	PPB	25 Sep 2023 14:34:44	U	SMPL
CCB56 - 1	CCB56	172	-0.0107	-	7470A	PPB	25 Sep 2023 14:36:59	U	SMPL
PB155755TB - 1	PB155755TB	529	0.0592	-	7470A	PPB	25 Sep 2023 14:39:14	U	SMPL
PB155757TB - 1	PB155757TB	545	0.0624	-	7470A	PPB	25 Sep 2023 14:41:30	U	SMPL
O4510-01LX5 - 1		119	-0.0210	-	7470A	PPB	25 Sep 2023 14:43:45	U	SMPL
O4510-01A - 1		19197	3.7151	-	7470A	PPB	25 Sep 2023 14:46:01	U	SMPL
O4563-03LX5 - 1		76	-0.0295	-	7470A	PPB	25 Sep 2023 14:48:16	U	SMPL
O4563-03A - 1		19120	3.7000	-	7470A	PPB	25 Sep 2023 14:50:32	U	SMPL
O4540-01LX5 - 1		84	-0.0279	-	7470A	PPB	25 Sep 2023 14:52:48	U	SMPL
O4540-01A - 1		20510	3.9722	-	7470A	PPB	25 Sep 2023 14:55:05	U	SMPL
CCV57 - 1	CCV57	25804	5.0089	-	7470A	PPB	25 Sep 2023 14:57:22	U	SMPL
CCB57 - 1	CCB57	173	-0.0105	-	7470A	PPB	25 Sep 2023 14:59:38	U	SMPL