



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127465

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/25/2023	17:20
2	CAL2	1	Water	NA	NA	20.5	6.99	09/25/2023	17:21
3	CAL3	1	Water	NA	NA	20.6	10.02	09/25/2023	17:22
4	ICV	1	Water	NA	NA	20.6	7.02	09/25/2023	17:25
5	CCV1	1	Water	NA	NA	20.6	2.01	09/25/2023	17:28
6	O4590-02	1	Water	NA	NA	20.7	6.44	09/25/2023	17:35
7	O4600-04	1	Water	NA	NA	20.3	7.49	09/25/2023	17:40
8	O4600-04DUP	1	Water	NA	NA	20.4	7.50	09/25/2023	17:41
9	CCV2	1	Water	NA	NA	20.6	12.02	09/25/2023	17:45

127465

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o4601 w WorkList ID : 173941 Department : Wet-Chemistry Date : 09-25-2023 17:14:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4590-02	F SAMPLE-1	Water	Corrosivity	Cool 4 deg C	SOUN03	N11	09/22/2023	9040C
O4600-04	A SOM-23-00088	Water	Corrosivity	Cool 4 deg C	PSEG03	E22	09/25/2023	9040C

Date/Time 09/25/23 17:20

Raw Sample Received by: SO(OC)

Raw Sample Relinquished by: ap sm

Date/Time 09/25/23

Raw Sample Received by: ap sm

Raw Sample Relinquished by: SO(OC)