

LB127471


7471B

INSTRUMENT 11: EX1

Linear

μ Abs.:

63653



A= 0.0000e+000

B= 1.5596e-004

C= -7.6887e-002

Rho= 0.9992244

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.0	0.000	-0.001	-0.001	489	0.000	489					-
0.2	0.200	0.190	-0.010	1712	0.0 %	1712					5
2.5	2.500	2.547	0.047	16823	0.0 %	16823					2
5.0	5.000	4.838	-0.162	31515	0.0 %	31515					-3
7.5	7.500	7.775	0.275	50348	0.0 %	50348					4
10.0	10.000	9.850	-0.150	63653	0.0 %	63653					-2

LB127471 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Sep 2023 11:09:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	489	-	0.0000	7471B	PPB	26 Sep 2023 11:24:00	S	Std
0.2 - 1	50.2	1712	-	0.2000	7471B	PPB	26 Sep 2023 11:26:17	S	Std
2.5 - 1	52.5	16823	-	2.5000	7471B	PPB	26 Sep 2023 11:28:35	S	Std
5.0 - 1	55	31515	-	5.0000	7471B	PPB	26 Sep 2023 11:30:52	S	Std
7.5 - 1	57.5	50348	-	7.5000	7471B	PPB	26 Sep 2023 11:35:59	S	Std
10.0 - 1	510	63653	-	10.0000	7471B	PPB	26 Sep 2023 11:38:14	S	Std
ICV80 - 1	ICV80	25435	3.8899	-	7471B	PPB	26 Sep 2023 11:41:03	U	SMPL
ICB80 - 1	ICB80	223	-0.0421	-	7471B	PPB	26 Sep 2023 11:43:18	U	SMPL
CCV58 - 1	CCV58	31145	4.7804	-	7471B	PPB	26 Sep 2023 11:45:34	U	SMPL
CCB58 - 1	CCB58	183	-0.0483	-	7471B	PPB	26 Sep 2023 11:47:49	U	SMPL
CRA - 1	CRA	1630	0.1773	-	7471B	PPB	26 Sep 2023 11:50:05	U	SMPL
HighStd - 1	HighStd	62971	9.7439	-	7471B	PPB	26 Sep 2023 11:52:20	U	SMPL
ChkStd - 1	ChkStd	44639	6.8849	-	7471B	PPB	26 Sep 2023 11:54:35	U	SMPL
PB155845BL - 1	PBS	-37	-0.0827	-	7471B	PPB	26 Sep 2023 11:56:53	U	SMPL
PB155845BS - 1	LCSS	27744	4.2500	-	7471B	PPB	26 Sep 2023 11:59:10	U	SMPL
O4566-02 - 1	CPT102-09	3777	0.5122	-	7471B	PPB	26 Sep 2023 12:01:26	U	SMPL
O4566-04 - 1	CPT102-06	1492	0.1558	-	7471B	PPB	26 Sep 2023 12:03:44	U	SMPL
O4566-06 - 1	B102-08-1	6347	0.9130	-	7471B	PPB	26 Sep 2023 12:06:00	U	SMPL
O4566-08 - 1	B102-08-1	12767	1.9142	-	7471B	PPB	26 Sep 2023 12:08:17	U	SMPL
O4566-10 - 1	B102-08-2	18234	2.7669	-	7471B	PPB	26 Sep 2023 12:10:32	U	SMPL
CCV59 - 1	CCV59	30819	4.7296	-	7471B	PPB	26 Sep 2023 12:12:47	U	SMPL
CCB59 - 1	CCB59	286	-0.0323	-	7471B	PPB	26 Sep 2023 12:15:03	U	SMPL
O4582-01 - 1	EO-02-092223	1979	0.2318	-	7471B	PPB	26 Sep 2023 12:22:33	U	SMPL
O4582-01DUP - 1	EO-02-092223DUP	2066	0.2453	-	7471B	PPB	26 Sep 2023 12:24:48	U	SMPL
O4582-01MS - 1	EO-02-092223MS	29545	4.5309	-	7471B	PPB	26 Sep 2023 12:27:04	U	SMPL
O4582-01MSD - 1	EO-02-092223MSD	31688	4.8651	-	7471B	PPB	26 Sep 2023 12:29:20	U	SMPL
O4587-01 - 1	SU-04-09222023	3458	0.4624	-	7471B	PPB	26 Sep 2023 12:31:35	U	SMPL
O4597-01 - 1	340	1733	0.1934	-	7471B	PPB	26 Sep 2023 12:33:53	U	SMPL
O4601-01 - 1	TP-1	1600	0.1726	-	7471B	PPB	26 Sep 2023 12:36:09	U	SMPL
O4601-04 - 1	TP-2	2100	0.2506	-	7471B	PPB	26 Sep 2023 12:38:26	U	SMPL
O4604-01 - 1	TR-04-092523	1634	0.1779	-	7471B	PPB	26 Sep 2023 12:40:42	U	SMPL
O4606-01 - 1	ETGI-327	10853	1.6157	-	7471B	PPB	26 Sep 2023 12:42:58	U	SMPL
CCV60 - 1	CCV60	30875	4.7383	-	7471B	PPB	26 Sep 2023 12:45:13	U	SMPL
CCB60 - 1	CCB60	333	-0.0250	-	7471B	PPB	26 Sep 2023 12:47:29	U	SMPL
O4582-01LX5 - 1		576	0.0129	-	7471B	PPB	26 Sep 2023 12:52:50	U	SMPL
O4582-01A - 1		30173	4.6288	-	7471B	PPB	26 Sep 2023 12:55:06	U	SMPL
CCV61 - 1	CCV61	31695	4.8662	-	7471B	PPB	26 Sep 2023 12:59:54	U	SMPL
CCB61 - 1	CCB61	395	-0.0153	-	7471B	PPB	26 Sep 2023 13:02:09	U	SMPL