



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127477

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/26/2023	11:00
2	O4580-05		1	12	100.00	>212.0	>212.0	09/26/2023	11:30
3	O4580-05DUP		1	12	100.00	>212.0	>212.0	09/26/2023	12:00
4	O4590-02		1	14	100.00	>212.0	>212.0	09/26/2023	12:30
5	O4600-04		1	14	100.00	>212.0	>212.0	09/26/2023	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6123477

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp 9-25

WorkList ID : 173895

Department : Wet-Chemistry

Date : 09-25-2023 08:09:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4580-05 D	ORANGE-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	N13	09/22/2023	1010B
O4600-04 B	SOM-23-00088	Water	Flash Point	Cool 4 deg C	PSEG03	E22	09/25/2023	1010B
O4590-02 E	SAMPLE-1	Water	Flash Point	Cool 4 deg C	SOUN03	N11	09/22/2023	1010B

Date/Time 09/26/2023 10:56
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: JH CWC

Date/Time 09/26/2023 13:46
Raw Sample Received by: JH CWC
Raw Sample Relinquished by: RM CWC