

L3127478


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

59547



A= 0.0000e+000

B= 1.6706e-004 slope

C= -6.4901e-002 y-intercept

Rho= 0.9997472

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	-0.032	-0.032	197	0.000	197					=
0.05	0.050					372	0				-12
0.20	0.200	0.177	-0.023	1445	0.0 %	1445					2
2.50	2.500	2.545	0.045	15621	0.0 %	15621					1
5.00	5.000	4.975	-0.025	30166	0.0 %	30166					2
7.50	7.500	7.653	0.153	46198	0.0 %	46198					-1
10.0	10.000	9.883	-0.117	59547	0.0 %	59547					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	197	-		0 SFAM01.1	PPB	9/26/2023 14:12	Std
	0.2 S01	1445	-		0.2 SFAM01.1	PPB	9/26/2023 14:17	Std
	2.5 S02	15621	-		2.5 SFAM01.1	PPB	9/26/2023 14:21	Std
	5 S03	30166	-		5 SFAM01.1	PPB	9/26/2023 14:24	Std
	7.5 S04	46198	-		7.5 SFAM01.1	PPB	9/26/2023 14:26	Std
	10 S05	59547	-		10 SFAM01.1	PPB	9/26/2023 14:31	Std
ICV081	ICV081	24538	4.0344 -		SFAM01.1	PPB	9/26/2023 14:34	SMPL
ICB081	ICB081	124	-0.0442 -		SFAM01.1	PPB	9/26/2023 14:36	SMPL
CCV061	CCV061	30162	4.974 -		SFAM01.1	PPB	9/26/2023 14:38	SMPL
CCB061	CCB061	123	-0.0444 -		SFAM01.1	PPB	9/26/2023 14:41	SMPL
PB155852BL	PBS852	221	-0.028 -		SFAM01.1	PPB	9/26/2023 14:43	SMPL
O4418-01	MEZYF0	2697	0.3857 -		SFAM01.1	PPB	9/26/2023 14:45	SMPL
O4418-02	MEZYF1	1979	0.2657 -		SFAM01.1	PPB	9/26/2023 14:47	SMPL
O4418-03	MEZYF2	566	0.0297 -		SFAM01.1	PPB	9/26/2023 14:50	SMPL
O4418-04	MEZYF3	1379	0.1655 -		SFAM01.1	PPB	9/26/2023 14:52	SMPL
O4418-05	MEZYF4	2461	0.3462 -		SFAM01.1	PPB	9/26/2023 14:54	SMPL
O4418-06	MEZYF5	7405	1.1722 -		SFAM01.1	PPB	9/26/2023 14:56	SMPL
O4418-08	MEZYM9	14288	2.3221 -		SFAM01.1	PPB	9/26/2023 15:06	SMPL
O4418-09	MEZYM9D	12453	2.0155 -		SFAM01.1	PPB	9/26/2023 15:08	SMPL
O4418-10	MEZYM9S	26475	4.358 -		SFAM01.1	PPB	9/26/2023 15:10	SMPL
O4418-11	MEZYN0	7746	1.2292 -		SFAM01.1	PPB	9/26/2023 15:13	SMPL
O4418-12	MEZYN1	6465	1.0151 -		SFAM01.1	PPB	9/26/2023 15:15	SMPL
O4418-13	MEZYN2	4583	0.7007 -		SFAM01.1	PPB	9/26/2023 15:17	SMPL
O4418-14	MEZYN3	4291	0.652 -		SFAM01.1	PPB	9/26/2023 15:19	SMPL
O4418-15	MEZYN4	8768	1.3999 -		SFAM01.1	PPB	9/26/2023 15:22	SMPL
O4418-16	MEZYX1	4868	0.7484 -		SFAM01.1	PPB	9/26/2023 15:24	SMPL
O4418-17	MEZYE4	1877	0.2487 -		SFAM01.1	PPB	9/26/2023 15:26	SMPL
O4418-18	MEZYE5	1785	0.2333 -		SFAM01.1	PPB	9/26/2023 15:29	SMPL
O4418-19	MEZYE6	3637	0.5427 -		SFAM01.1	PPB	9/26/2023 15:31	SMPL
O4418-20	MEZYE7	2257	0.3122 -		SFAM01.1	PPB	9/26/2023 15:33	SMPL
CCV062	CCV062	29676	4.8928 -		SFAM01.1	PPB	9/26/2023 15:35	SMPL
CCB062	CCB062	157	-0.0387 -		SFAM01.1	PPB	9/26/2023 15:38	SMPL
O4418-21	MEZYE8	4448	0.6782 -		SFAM01.1	PPB	9/26/2023 15:40	SMPL
O4418-22	MEZYE9	11059	1.7826 -		SFAM01.1	PPB	9/26/2023 15:42	SMPL
CCV063	CCV063	29677	4.893 -		SFAM01.1	PPB	9/26/2023 15:44	SMPL
CCB063	CCB063	161	-0.038 -		SFAM01.1	PPB	9/26/2023 15:47	SMPL