

LB 127491


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45930



A= 0.0000e+000

B= 2.1567e-004

C= -6.9421e-002

Rho= 0.9995427

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.021	-0.021	225	0.000	225	0				-5
0.2	0.200	0.190	-0.010	1204	0.0 %	1204					-1
2.5	2.500	2.480	-0.020	11820	0.0 %	11820					0
5.0	5.000	4.998	-0.002	23495	0.0 %	23495					3
7.5	7.500	7.717	0.217	36102	0.0 %	36102					-2
10.0	10.000	9.836	-0.164	45930	0.0 %	45930					

LB127491 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Sep 2023 12:34:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	225	-	0.0000	7470A	PPB	27 Sep 2023 13:14:15	S	Std
0.2 - 1	50.2	1204	-	0.2000	7470A	PPB	27 Sep 2023 13:16:32	S	Std
2.5 - 1	52.5	11820	-	2.5000	7470A	PPB	27 Sep 2023 13:18:49	S	Std
5.0 - 1	55	23495	-	5.0000	7470A	PPB	27 Sep 2023 13:21:06	S	Std
7.5 - 1	57.5	36102	-	7.5000	7470A	PPB	27 Sep 2023 13:23:22	S	Std
10.0 - 1	310	45930	-	10.0000	7470A	PPB	27 Sep 2023 13:28:40	S	Std
ICV82 - 1	ICV82	18848	3.9955	-	7470A	PPB	27 Sep 2023 13:32:35	U	SMPL
ICB82 - 1	ICB82	224	-0.0211	-	7470A	PPB	27 Sep 2023 13:34:50	U	SMPL
CCV64 - 1	CCV64	23494	4.9976	-	7470A	PPB	27 Sep 2023 13:39:36	U	SMPL
CCB64 - 1	CCB64	245	-0.0166	-	7470A	PPB	27 Sep 2023 13:41:51	U	SMPL
CRA - 1	CRA	1412	0.2351	-	7470A	PPB	27 Sep 2023 13:50:09	U	SMPL
HighStd - 1	HighStd	45936	9.8376	-	7470A	PPB	27 Sep 2023 13:52:24	U	SMPL
ChkStd - 1	ChkStd	30712	6.5543	-	7470A	PPB	27 Sep 2023 13:54:39	U	SMPL
PB155850BL - 1	PBW	-54	-0.0811	-	7470A	PPB	27 Sep 2023 13:56:55	U	SMPL
PB155850BS - 1	LCSW	19289	4.0907	-	7470A	PPB	27 Sep 2023 13:59:12	U	SMPL
O4597-02 - 1	340	1	-0.0692	-	7470A	PPB	27 Sep 2023 14:01:28	U	SMPL
O4600-01 - 1	COMP-1	68	-0.0548	-	7470A	PPB	27 Sep 2023 14:03:44	U	SMPL
O4600-02 - 1	COMP-2	53	-0.0580	-	7470A	PPB	27 Sep 2023 14:06:01	U	SMPL
O4600-03 - 1	COMP-3	80	-0.0522	-	7470A	PPB	27 Sep 2023 14:08:18	U	SMPL
O4601-02 - 1	TP-1	94	-0.0491	-	7470A	PPB	27 Sep 2023 14:10:33	U	SMPL
CCV65 - 1	CCV65	24100	5.1282	-	7470A	PPB	27 Sep 2023 14:12:48	U	SMPL
CCB65 - 1	CCB65	265	-0.0123	-	7470A	PPB	27 Sep 2023 14:15:02	U	SMPL
O4601-05 - 1	TP-2	104	-0.0470	-	7470A	PPB	27 Sep 2023 14:17:17	U	SMPL
O4601-05DUP - 1	TP-2DUP	269	-0.0114	-	7470A	PPB	27 Sep 2023 14:19:33	U	SMPL
O4601-05MS - 1	TP-2MS	16811	3.5562	-	7470A	PPB	27 Sep 2023 14:21:48	U	SMPL
O4601-05MSD - 1	TP-2MSD	16346	3.4559	-	7470A	PPB	27 Sep 2023 14:24:03	U	SMPL
PB155851BL - 1	PBW	-8	-0.0711	-	7470A	PPB	27 Sep 2023 14:26:19	U	SMPL
PB155851BS - 1	LCSW	19246	4.0814	-	7470A	PPB	27 Sep 2023 14:28:35	U	SMPL
O4590-02 - 1	SAMPLE-1	41	-0.0606	-	7470A	PPB	27 Sep 2023 14:30:51	U	SMPL
O4600-04 - 1	SOM-23-00088	117	-0.0442	-	7470A	PPB	27 Sep 2023 14:33:08	U	SMPL
O4600-04DUP - 1	SOM-23-00088DUP	135	-0.0403	-	7470A	PPB	27 Sep 2023 14:35:24	U	SMPL
O4600-04MS - 1	SOM-23-00088MS	16964	3.5892	-	7470A	PPB	27 Sep 2023 14:37:40	U	SMPL
CCV66 - 1	CCV66	24609	5.2380	-	7470A	PPB	27 Sep 2023 14:44:44	U	SMPL
CCB66 - 1	CCB66	203	-0.0256	-	7470A	PPB	27 Sep 2023 14:46:59	U	SMPL
PB155810TB - 1	PB155810TB	-79	-0.0865	-	7470A	PPB	27 Sep 2023 14:51:30	U	SMPL
PB155813TB - 1	PB155813TB	57	-0.0571	-	7470A	PPB	27 Sep 2023 14:53:46	U	SMPL
O4600-4MSD - 1	SOM-23-00088MSD	17932	3.7980	-	7470A	PPB	27 Sep 2023 14:56:01	U	SMPL
O4601-05LX5 - 1		-48	-0.0798	-	7470A	PPB	27 Sep 2023 14:58:17	U	SMPL
O4601-05A - 1		18948	4.0171	-	7470A	PPB	27 Sep 2023 15:00:33	U	SMPL
O4600-04LX5 - 1		17	-0.0658	-	7470A	PPB	27 Sep 2023 15:02:49	U	SMPL
O4600-04A - 1		19389	4.1122	-	7470A	PPB	27 Sep 2023 15:05:07	U	SMPL
CCV67 - 1	CCV67	24623	5.2410	-	7470A	PPB	27 Sep 2023 15:07:23	U	SMPL
CCB67 - 1	CCB67	242	-0.0172	-	7470A	PPB	27 Sep 2023 15:09:39	U	SMPL