



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127506

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/27/2023	16:10
2	CAL2	1	Water	NA	NA	20.6	7.00	09/27/2023	16:11
3	CAL3	1	Water	NA	NA	20.6	10.03	09/27/2023	16:15
4	ICV	1	Water	NA	NA	20.5	7.00	09/27/2023	16:20
5	CCV1	1	Water	NA	NA	20.7	2.01	09/27/2023	17:10
6	O4641-03	1	Water	NA	NA	20.4	6.13	09/27/2023	17:15
7	O4641-03DUP	1	Water	NA	NA	20.5	6.14	09/27/2023	17:16
8	O4643-02	1	Water	NA	NA	20.1	7.53	09/27/2023	17:20
9	CCV2	1	Water	NA	NA	20.6	12.02	09/27/2023	17:25

WORKLIST(Hardcopy Internal Chain)

12506

WorkList Name : corrosivity 04620 water WorkList ID : 174025 Department : Wet-Chemistry Date : 09-27-2023 17:01:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4641-03	BS-RAINWATER	Water	Corrosivity	Cool 4 deg C	PSEG03	E32	09/27/2023	9040C
O4643-02	217	Water	Corrosivity	Cool 4 deg C	PSEG03	E32	09/27/2023	9040C

Date/Time 09/17/13 16:00
 Raw Sample Received by: SA (WC)
 Raw Sample Relinquished by: A.C. Smith

Date/Time 09/17/13 18:00
 Raw Sample Received by: J. G. Smith
 Raw Sample Relinquished by: J. G. Smith