

L3127512


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

45113



A= 0.0000e+000

B= 2.2554e-004

C= -1.0937e-001

Rho= 0.9998601

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01)
0.0	0.000	0.009	0.009	523	0.000	523					6
0.05	0.050	0.053	0.003	719	0.0 %	719	0				5
0.2	0.200	0.210	0.010	1416	0.0 %	1416					1
2.5	2.500	2.471	-0.029	11441	0.0 %	11441					1
5.0	5.000	5.071	0.071	22968	0.0 %	22968					1
7.5	7.500	7.371	-0.129	33166	0.0 %	33166					2
10.0	10.000	10.066	0.066	45113	0.0 %	45113					1

LB127512 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Sep 2023 11:19:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	523	-	0.0000	245.1	PPB	28 Sep 2023 11:33:38	S	Std
0.05 - 1	50.05	719	-	0.0500	245.1	PPB	28 Sep 2023 11:35:55	S	Std
0.2 - 1	50.2	1416	-	0.2000	245.1	PPB	28 Sep 2023 11:38:12	S	Std
2.5 - 1	52.5	11441	-	2.5000	245.1	PPB	28 Sep 2023 11:40:29	S	Std
5.0 - 1	55	22968	-	5.0000	245.1	PPB	28 Sep 2023 11:42:45	S	Std
7.5 - 1	57.5	33166	-	7.5000	245.1	PPB	28 Sep 2023 11:45:01	S	Std
10.0 - 1	510	45113	-	10.0000	245.1	PPB	28 Sep 2023 11:55:15	S	Std
ICV84 - 1	ICV84	17462	3.8291	-245.1		PPB	28 Sep 2023 11:58:00	U	SMPL
ICB84 - 1	ICB84	342	-0.0322	-245.1		PPB	28 Sep 2023 12:00:14	U	SMPL
CCV73 - 1	CCV73	22869	5.0486	-245.1		PPB	28 Sep 2023 12:02:30	U	SMPL
CCB73 - 1	CCB73	358	-0.0286	-245.1		PPB	28 Sep 2023 12:04:44	U	SMPL
CRA - 1	CRA	695	0.0474	-245.1		PPB	28 Sep 2023 12:09:45	U	SMPL
HighStd - 1	HighStd	46633	10.4084	-245.1		PPB	28 Sep 2023 12:11:59	U	SMPL
ChkStd - 1	ChkStd	30642	6.8017	-245.1		PPB	28 Sep 2023 12:14:14	U	SMPL
PB155930BL - 1	PBW	224	-0.0588	-245.1		PPB	28 Sep 2023 12:16:32	U	SMPL
PB155930BS - 1	LCSW	17807	3.9069	-245.1		PPB	28 Sep 2023 12:21:26	U	SMPL
O4590-01 - 1	SAMPLE-1	3141	0.5991	-245.1		PPB	28 Sep 2023 12:28:15	U	SMPL
O4590-01DUP - 1	SAMPLE-1DUP	2989	0.5648	-245.1		PPB	28 Sep 2023 12:30:31	U	SMPL
O4590-01MS - 1	SAMPLE-1MS	22533	4.9728	-245.1		PPB	28 Sep 2023 12:35:31	U	SMPL
O4590-01MSD - 1	SAMPLE-1MSD	20698	4.5589	-245.1		PPB	28 Sep 2023 12:37:47	U	SMPL
CCV74 - 1	CCV74	22957	5.0684	-245.1		PPB	28 Sep 2023 12:40:04	U	SMPL
CCB74 - 1	CCB74	459	-0.0058	-245.1		PPB	28 Sep 2023 12:42:20	U	SMPL
O4590-01LX5 - 1		1005	0.1173	-245.1		PPB	28 Sep 2023 12:44:36	U	SMPL
O4590-01A - 1		20156	4.4367	-245.1		PPB	28 Sep 2023 12:46:50	U	SMPL
CCV75 - 1	CCV75	22746	5.0209	-245.1		PPB	28 Sep 2023 12:49:05	U	SMPL
CCB75 - 1	CCB75	451	-0.0076	-245.1		PPB	28 Sep 2023 12:51:21	U	SMPL