

LB127517


7471B

INSTRUMENT 1A: CV1

Linear

μ Abs.:

66480



A= 0.0000e+000

B= 1.5197e-004

C= -5.7624e-002

Rho= 0.9999525

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0/0
0.0	0.000	0.015	0.015	476	0.000	476					4
0.2	0.200	0.209	0.009	1756	0.0 %	1756					1
2.5	2.500	2.522	0.022	16972	0.0 %	16972					1
5.0	5.000	4.934	-0.066	32847	0.0 %	32847					0
7.5	7.500	7.475	-0.025	49569	0.0 %	49569					0
10.0	10.000	10.045	0.045	66480	0.0 %	66480					0

LB127517 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Sep 2023 13:52:52

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	476	-	0.0000	7471B	PPB	28 Sep 2023 14:06:49	S	Std
0.2 - 1	51.2	1756	-	0.2000	7471B	PPB	28 Sep 2023 14:09:06	S	Std
2.5 - 1	52.5	16972	-	2.5000	7471B	PPB	28 Sep 2023 14:11:23	S	Std
5.0 - 1	55	32847	-	5.0000	7471B	PPB	28 Sep 2023 14:13:39	S	Std
7.5 - 1	57.5	49569	-	7.5000	7471B	PPB	28 Sep 2023 14:18:35	S	Std
10.0 - 1	510	66480	-	10.0000	7471B	PPB	28 Sep 2023 14:20:51	S	Std
ICV85 - 1	ICV85	26837	4.0207	-	7471B	PPB	28 Sep 2023 14:23:40	U	SMPL
ICB85 - 1	ICB85	298	-0.0123	-	7471B	PPB	28 Sep 2023 14:25:55	U	SMPL
CCV76 - 1	CCV76	33322	5.0062	-	7471B	PPB	28 Sep 2023 14:28:10	U	SMPL
CCB76 - 1	CCB76	309	-0.0107	-	7471B	PPB	28 Sep 2023 14:30:25	U	SMPL
CRA - 1	CRA	1863	0.2255	-	7471B	PPB	28 Sep 2023 14:32:40	U	SMPL
HighStd - 1	HighStd	66036	9.9777	-	7471B	PPB	28 Sep 2023 14:34:54	U	SMPL
ChkStd - 1	ChkStd	47867	7.2166	-	7471B	PPB	28 Sep 2023 14:37:09	U	SMPL
PB155941BL - 1	PBS	-73	-0.0687	-	7471B	PPB	28 Sep 2023 14:39:27	U	SMPL
PB155941BS - 1	LCSS	26810	4.0166	-	7471B	PPB	28 Sep 2023 14:44:14	U	SMPL
O4614-01 - 1	AUD-1053	874	0.0752	-	7471B	PPB	28 Sep 2023 14:46:30	U	SMPL
O4614-02 - 1	AUD-1055	13107	1.9342	-	7471B	PPB	28 Sep 2023 14:48:46	U	SMPL
O4614-03 - 1	AUD-1048	263	-0.0177	-	7471B	PPB	28 Sep 2023 14:51:03	U	SMPL
O4614-04 - 1	AUD-1049	517	0.0209	-	7471B	PPB	28 Sep 2023 14:53:19	U	SMPL
O4620-01 - 1	CON-PAD	374	-0.0008	-	7471B	PPB	28 Sep 2023 14:55:34	U	SMPL
CCV77 - 1	CCV77	33269	4.9982	-	7471B	PPB	28 Sep 2023 14:57:48	U	SMPL
CCB77 - 1	CCB77	366	-0.0020	-	7471B	PPB	28 Sep 2023 15:00:05	U	SMPL
O4620-03 - 1	CON-WALL	369	-0.0015	-	7471B	PPB	28 Sep 2023 15:02:19	U	SMPL
O4620-05 - 1	CON-SUPPORT-A	469	0.0136	-	7471B	PPB	28 Sep 2023 15:04:34	U	SMPL
O4620-07 - 1	CON-SUPPORT-B	343	-0.0055	-	7471B	PPB	28 Sep 2023 15:06:50	U	SMPL
O4620-09 - 1	STONE-IN	279	-0.0152	-	7471B	PPB	28 Sep 2023 15:09:06	U	SMPL
O4620-12 - 1	STONE-OUT	191	-0.0286	-	7471B	PPB	28 Sep 2023 15:11:21	U	SMPL
O4621-02 - 1	ETGI-287	4512	0.6281	-	7471B	PPB	28 Sep 2023 15:13:37	U	SMPL
O4622-01 - 1	I-1(0-5)	14672	2.1720	-	7471B	PPB	28 Sep 2023 15:15:53	U	SMPL
O4622-02 - 1	I-4(0-5)	79323	11.9969	-	7471B	PPB	28 Sep 2023 15:18:10	U	SMPL
O4622-02DUP - 1	I-4(0-5)DUP	73361	11.0909	-	7471B	PPB	28 Sep 2023 15:20:27	U	SMPL
O4622-02MS - 1	I-4(0-5)MS	85748	12.9733	-	7471B	PPB	28 Sep 2023 15:26:56	U	SMPL
CCV78 - 1	CCV78	33226	4.9916	-	7471B	PPB	28 Sep 2023 15:29:14	U	SMPL
CCB78 - 1	CCB78	306	-0.0111	-	7471B	PPB	28 Sep 2023 15:31:31	U	SMPL
O4622-02MSD - 1	I-4(0-5)MSD	95412	14.4419	-	7471B	PPB	28 Sep 2023 15:33:46	U	SMPL
PB155942BL - 1	PBS	-7	-0.0587	-	7471B	PPB	28 Sep 2023 15:36:01	U	SMPL
O4627-01 - 1	280-300FB-BACKFILL-01	647	0.0407	-	7471B	PPB	28 Sep 2023 15:40:34	U	SMPL
O4627-01DUP - 1	280-300FB-BACKFILL-01DUP	346	-0.0050	-	7471B	PPB	28 Sep 2023 15:42:50	U	SMPL
O4627-01MS - 1	280-300FB-BACKFILL-01MS	27024	4.0491	-	7471B	PPB	28 Sep 2023 15:45:05	U	SMPL
O4627-01MSD - 1	280-300FB-BACKFILL-01MSD	27126	4.0646	-	7471B	PPB	28 Sep 2023 15:47:22	U	SMPL
O4627-03 - 1	280-300FB-BACKFILL-02	583	0.0310	-	7471B	PPB	28 Sep 2023 15:49:39	U	SMPL
O4627-05 - 1	280-300FB-BACKFILL-03	325	-0.0082	-	7471B	PPB	28 Sep 2023 15:51:55	U	SMPL
O4627-07 - 1	280-300FB-BACKFILL-04	274	-0.0160	-	7471B	PPB	28 Sep 2023 15:54:11	U	SMPL
CCV79 - 1	CCV79	33379	5.0149	-	7471B	PPB	28 Sep 2023 15:56:26	U	SMPL
CCB79 - 1	CCB79	351	-0.0043	-	7471B	PPB	28 Sep 2023 15:58:42	U	SMPL
O4627-09 - 1	280-300FB-BACKFILL-05	411	0.0048	-	7471B	PPB	28 Sep 2023 16:00:57	U	SMPL
O4627-11 - 1	280-300FB-BACKFILL-06	251	-0.0195	-	7471B	PPB	28 Sep 2023 16:03:12	U	SMPL
O4627-13 - 1	280-300FB-BACKFILL-07	249	-0.0198	-	7471B	PPB	28 Sep 2023 16:05:28	U	SMPL
O4627-15 - 1	280-300FB-BACKFILL-08	269	-0.0167	-	7471B	PPB	28 Sep 2023 16:07:43	U	SMPL
O4627-17 - 1	280-300FB-BACKFILL-09	222	-0.0239	-	7471B	PPB	28 Sep 2023 16:09:59	U	SMPL
O4627-19 - 1	280-300FB-BACKFILL-10	236	-0.0218	-	7471B	PPB	28 Sep 2023 16:12:15	U	SMPL
O4627-21 - 1	280-300FB-BACKFILL-11	208	-0.0260	-	7471B	PPB	28 Sep 2023 16:14:31	U	SMPL
O4627-23 - 1	280-300FB-BACKFILL-12	192	-0.0284	-	7471B	PPB	28 Sep 2023 16:16:46	U	SMPL
O4627-25 - 1	280-300FB-BACKFILL-13	193	-0.0283	-	7471B	PPB	28 Sep 2023 16:19:05	U	SMPL
O4627-27 - 1	280-300FB-BACKFILL-14	627	0.0377	-	7471B	PPB	28 Sep 2023 16:21:21	U	SMPL
CCV80 - 1	CCV80	0	-0.0576	-	7471B	PPB	28 Sep 2023 16:23:37	U	SMPL
CCB80 - 1	CCB80	33209	4.9891	-	7471B	PPB	28 Sep 2023 16:29:44	U	SMPL
O4636-01 - 1	OR-03-09272023	2689	0.3510	-	7471B	PPB	28 Sep 2023 16:32:00	U	SMPL
O4637-01 - 1	WC-1	22971	3.4332	-	7471B	PPB	28 Sep 2023 16:34:15	U	SMPL
O4642-01 - 1	PAT-01-09272023	9723	1.4200	-	7471B	PPB	28 Sep 2023 16:36:30	U	SMPL
O4622-02LX5 - 1		18765	2.7940	-	7471B	PPB	28 Sep 2023 16:38:45	U	SMPL
O4622-02A - 1		100444	15.2066	-	7471B	PPB	28 Sep 2023 16:41:01	U	SMPL
O4627-01LX5 - 1		-57	-0.0663	-	7471B	PPB	28 Sep 2023 16:43:17	U	SMPL
O4627-01A - 1		27406	4.1072	-	7471B	PPB	28 Sep 2023 16:45:33	U	SMPL
PB155942BS - 1	LCSS	26578	3.9814	-	7471B	PPB	28 Sep 2023 16:47:50	U	SMPL
CCV81 - 1	CCV81	32808	4.9281	-	7471B	PPB	28 Sep 2023 16:50:08	U	SMPL
CCB81 - 1	CCB81	330	-0.0075	-	7471B	PPB	28 Sep 2023 16:52:25	U	SMPL
O4622-02DLX2 - 1		37114	5.5825	-	7471B	PPB	28 Sep 2023 16:54:41	U	SMPL
O4622-02DUPDLX2 - 1		38170	5.7430	-	7471B	PPB	28 Sep 2023 16:56:57	U	SMPL
O4622-02MSDLX2 - 1		48296	7.2818	-	7471B	PPB	28 Sep 2023 16:59:13	U	SMPL
O4622-02MSDDLX2 - 1		51712	7.8009	-	7471B	PPB	28 Sep 2023 17:01:28	U	SMPL
O4622-02LDLX10 - 1		7426	1.0709	-	7471B	PPB	28 Sep 2023 17:03:45	U	SMPL
O4622-02ADLX2 - 1		49511	7.4664	-	7471B	PPB	28 Sep 2023 17:06:01	U	SMPL
CCV82 - 1	CCV82	33633	5.0535	-	7471B	PPB	28 Sep 2023 17:08:17	U	SMPL
CCB82 - 1	CCB82	348	-0.0047	-	7471B	PPB	28 Sep 2023 17:10:33	U	SMPL
							28 Sep 2023 17:12:50	U	SMPL