



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127519

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/28/2023	12:15
2	CAL2	1	Water	NA	NA	20.6	7.00	09/28/2023	12:16
3	CAL3	1	Water	NA	NA	20.7	10.03	09/28/2023	12:19
4	ICV	1	Water	NA	NA	20.5	7.01	09/28/2023	12:20
5	CCV1	1	Water	NA	NA	20.5	2.01	09/28/2023	12:25
6	O4661-02	1	Water	NA	NA	20.4	5.62	09/28/2023	12:30
7	O4661-02DUP	1	Water	NA	NA	20.5	5.63	09/28/2023	12:31
8	O4661-03	1	Water	NA	NA	20.4	6.18	09/28/2023	12:37
9	O4661-04	1	Water	NA	NA	20.3	6.25	09/28/2023	12:45
10	O4661-05	1	Water	NA	NA	20.6	6.21	09/28/2023	12:48
11	O4665-01	1	Water	NA	NA	20.1	5.06	09/28/2023	12:50
12	CCV2	1	Water	NA	NA	20.5	12.02	09/28/2023	13:00

WORKLIST(Hardcopy Internal Chain)

UB 127619

WorkList Name : ph 04661

WorkList ID : 174062

Department : Wet-Chemistry

Date : 09-28-2023 11:55:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4661-02	TT-054-IDWGW-20230927	Water	pH	Cool 4 deg C	TETR06	E31	09/27/2023	9040C
O4661-03	TT-055-IDWGW-20230927	Water	pH	Cool 4 deg C	TETR06	E31	09/27/2023	9040C
O4661-04	TT-056-IDWGW-20230927	Water	pH	Cool 4 deg C	TETR06	E31	09/27/2023	9040C
O4661-05	TT-057-IDWGW-20230927	Water	pH	Cool 4 deg C	TETR06	E31	09/27/2023	9040C
O4665-01	POOL-WATER	Water	pH	Cool 4 deg C	STER06	E52	09/26/2023	9040C

Date/Time 09/28/23 12:10

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time

09/28/23 16:00

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC