



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB127523
Thermometer ID: Flashpoint

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 771.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/28/2023	13:20
2	O4596-01		1	12	100.00	>212.0	>212.0	09/28/2023	13:50
3	O4613-01		1	11	100.00	>212.0	>212.0	09/28/2023	14:20
4	O4613-02		1	12	100.00	>212.0	>212.0	09/28/2023	14:50
5	O4613-03		1	14	100.00	>212.0	>212.0	09/28/2023	15:20
6	O4614-05		1	11	100.00	>212.0	>212.0	09/28/2023	15:50
7	O4614-05DUP		1	11	100.00	>212.0	>212.0	09/28/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127523

WorkList Name : fp s 9-26

WorkList ID : 173956

Department : Wet-Chemistry

Date : 09-26-2023 08:41:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4596-01	0817	Solid	Flash Point	Cool 4 deg C	PSEG03	E21	09/25/2023	1010B
O4613-01	1029	Solid	Flash Point	Cool 4 deg C	PSEG03	E21	09/26/2023	1010B
O4613-02	1030	Solid	Flash Point	Cool 4 deg C	PSEG03	E21	09/26/2023	1010B
O4613-03	1031	Solid	Flash Point	Cool 4 deg C	PSEG03	E21	09/26/2023	1010B
O4614-05	AUD-1056	Solid	Flash Point	Cool 4 deg C	PSEG03	E22	09/26/2023	1010B

Date/Time 09/28/2023 13:06

Raw Sample Received by: RA CWL

Raw Sample Relinquished by: 78 (CWC)

Date/Time 09/28/2023 16:30
Raw Sample Received by: 78 (CWC)
Raw Sample Relinquished by: RA CWL