



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127536

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/29/2023	08:15
2	CAL2	1	Water	NA	NA	20.5	7.00	09/29/2023	08:16
3	CAL3	1	Water	NA	NA	20.6	10.02	09/29/2023	08:19
4	ICV	1	Water	NA	NA	20.6	7.01	09/29/2023	08:20
5	CCV1	1	Water	NA	NA	20.8	2.01	09/29/2023	08:22
6	O4651-03	1	Solid	20.02	20	20.7	8.90	09/29/2023	08:25
7	O4651-03DUP	1	Solid	20.03	20	20.8	8.91	09/29/2023	08:26
8	O4652-03	1	Solid	20.02	20	20.1	8.32	09/29/2023	08:30
9	O4652-06	1	Solid	20.01	20	20.4	8.29	09/29/2023	08:35
10	O4652-09	1	Solid	20.02	20	20.4	8.36	09/29/2023	08:39
11	O4652-12	1	Solid	20.03	20	20.3	8.31	09/29/2023	08:42
12	O4652-15	1	Solid	20.02	20	20.2	8.34	09/29/2023	08:45
13	O4652-18	1	Solid	20.03	20	20.2	8.34	09/29/2023	08:50
14	CCV2	1	Water	NA	NA	20.6	12.02	09/29/2023	08:55

WORKLIST(Hardcopy Internal Chain)

127536

WorkList Name : ph 04651

WorkList ID : 174075

Department : Wet-Chemistry

Date : 09-29-2023 07:55:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4651-03	SP1-COMP	Solid	pH	Cool 4 deg C	EART12	E41	09/27/2023	9045D
O4652-03	IS1-COMP	Solid	pH	Cool 4 deg C	EART12	D31	09/27/2023	9045D
O4652-06	IS2-COMP	Solid	pH	Cool 4 deg C	EART12	D31	09/27/2023	9045D
O4652-09	IS3-COMP	Solid	pH	Cool 4 deg C	EART12	D31	09/27/2023	9045D
O4652-12	IS4-COMP	Solid	pH	Cool 4 deg C	EART12	D31	09/27/2023	9045D
O4652-15	IS5-COMP	Solid	pH	Cool 4 deg C	EART12	D31	09/27/2023	9045D
O4652-18	IS6-COMP	Solid	pH	Cool 4 deg C	EART12	D31	09/27/2023	9045D

Date/Time

09/29/23 08:10

Raw Sample Received by:

201911

Raw Sample Relinquished by:

201911

Date/Time

09/29/23

Raw Sample Received by:

201911

Raw Sample Relinquished by:

201911