


LB127538

7470A

INSTRUMENT ID: CV1

 μ Abs.:

1 47275



Linear

A= 0.0000e+000

B= 2.1154e-004

C= -1.6077e-001

Rho= 0.9996159

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	2SD
0.0	0.000	-0.054	-0.054	505	0.000	505	0				2
0.2	0.200	0.205	0.005	1727	0.0 %	1727					-2
2.5	2.500	2.461	-0.039	12392	0.0 %	12392					2
5.0	5.000	5.093	0.093	24836	0.0 %	24836					2
7.5	7.500	7.655	0.155	36950	0.0 %	36950					2
10.0	10.000	9.840	-0.160	47279	0.0 %	47279					-2

LB127538 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Sep 2023 09:52:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	505	-	0.0000	7470A	PPB	29 Sep 2023 09:58:36	S	Std
0.2 - 1	50.2	1727	-	0.2000	7470A	PPB	29 Sep 2023 10:00:53	S	Std
2.5 - 1	52.5	12392	-	2.5000	7470A	PPB	29 Sep 2023 10:03:10	S	Std
5.0 - 1	55	24836	-	5.0000	7470A	PPB	29 Sep 2023 10:05:27	S	Std
7.5 - 1	57.5	36950	-	7.5000	7470A	PPB	29 Sep 2023 10:07:43	S	Std
10.0 - 1	510	47279	-	10.0000	7470A	PPB	29 Sep 2023 10:12:36	S	Std
ICV86 - 1	ICV86	18810	3.8182	-	7470A	PPB	29 Sep 2023 10:15:24	U	SMPL
ICB86 - 1	ICB86	938	0.0376	-	7470A	PPB	29 Sep 2023 10:17:38	U	SMPL
CCV83 - 1	CCV83	22539	4.6070	-	7470A	PPB	29 Sep 2023 10:19:53	U	SMPL
CCB83 - 1	CCB83	966	0.0436	-	7470A	PPB	29 Sep 2023 10:22:08	U	SMPL
CRA - 1	CRA	1777	0.2151	-	7470A	PPB	29 Sep 2023 10:30:55	U	SMPL
HighStd - 1	HighStd	47376	9.8609	-	7470A	PPB	29 Sep 2023 10:33:10	U	SMPL
ChkStd - 1	ChkStd	32939	6.8070	-	7470A	PPB	29 Sep 2023 10:35:25	U	SMPL
PB155935BL - 1	PBW	296	-0.0982	-	7470A	PPB	29 Sep 2023 10:37:42	U	SMPL
PB155935BS - 1	LCSW	21235	4.3312	-	7470A	PPB	29 Sep 2023 10:39:58	U	SMPL
O4620-02 - 1	CON-PAD	370	-0.0825	-	7470A	PPB	29 Sep 2023 10:42:14	U	SMPL
O4620-04 - 1	CON-WALL	445	-0.0666	-	7470A	PPB	29 Sep 2023 10:44:30	U	SMPL
O4620-06 - 1	CON-SUPPORT-A	433	-0.0692	-	7470A	PPB	29 Sep 2023 10:46:47	U	SMPL
O4620-08 - 1	CON-SUPPORT-B	493	-0.0565	-	7470A	PPB	29 Sep 2023 10:49:04	U	SMPL
O4620-10 - 1	STONE-IN	484	-0.0584	-	7470A	PPB	29 Sep 2023 10:51:19	U	SMPL
CCV84 - 1	CCV84	22859	4.6747	-	7470A	PPB	29 Sep 2023 10:53:34	U	SMPL
CCB84 - 1	CCB84	1002	0.0512	-	7470A	PPB	29 Sep 2023 10:55:49	U	SMPL
O4620-13 - 1	STONE-OUT	425	-0.0709	-	7470A	PPB	29 Sep 2023 10:58:05	U	SMPL
O4641-01 - 1	BS-OIL-DRUM	483	-0.0586	-	7470A	PPB	29 Sep 2023 11:00:20	U	SMPL
O4641-02 - 1	BS-OIL-TOTE	402	-0.0757	-	7470A	PPB	29 Sep 2023 11:02:35	U	SMPL
O4643-01 - 1	215	426	-0.0707	-	7470A	PPB	29 Sep 2023 11:04:50	U	SMPL
O4648-01 - 1	D3242	85	-0.1428	-	7470A	PPB	29 Sep 2023 11:13:25	U	SMPL
O4648-01DUP - 1	D3242DUP	60	-0.1481	-	7470A	PPB	29 Sep 2023 11:15:40	U	SMPL
O4648-01MS - 1	D3242MS	15030	3.0186	-	7470A	PPB	29 Sep 2023 11:17:57	U	SMPL
O4648-01MSD - 1	D3242MSD	13506	2.6962	-	7470A	PPB	29 Sep 2023 11:22:44	U	SMPL
PB155936BL - 1	PBW	466	-0.0622	-	7470A	PPB	29 Sep 2023 11:25:00	U	SMPL
CCV85 - 1	CCV85	25335	5.1985	-	7470A	PPB	29 Sep 2023 11:33:20	U	SMPL
CCB85 - 1	CCB85	595	-0.0349	-	7470A	PPB	29 Sep 2023 11:35:35	U	SMPL
PB155936BS - 1	LCSW	17421	3.5244	-	7470A	PPB	29 Sep 2023 11:37:51	U	SMPL
O4641-03 - 1	BS-RAINWATER	332	-0.0905	-	7470A	PPB	29 Sep 2023 11:40:06	U	SMPL
O4641-03DUP - 1	BS-RAINWATERDUP	379	-0.0806	-	7470A	PPB	29 Sep 2023 11:42:22	U	SMPL
O4641-03MS - 1	BS-RAINWATERMS	16050	3.2344	-	7470A	PPB	29 Sep 2023 11:44:37	U	SMPL
O4641-03MSD - 1	BS-RAINWATERMSD	16705	3.3729	-	7470A	PPB	29 Sep 2023 11:46:53	U	SMPL
O4643-02 - 1	217	387	-0.0789	-	7470A	PPB	29 Sep 2023 11:49:09	U	SMPL
CCV86 - 1	CCV86	24145	4.9468	-	7470A	PPB	29 Sep 2023 11:51:26	U	SMPL
CCB86 - 1	CCB86	384	-0.0795	-	7470A	PPB	29 Sep 2023 11:53:43	U	SMPL
PB155882TB - 1	PB155882TB	399	-0.0764	-	7470A	PPB	29 Sep 2023 11:56:00	U	SMPL
PB155927TB - 1	PB155927TB	444	-0.0669	-	7470A	PPB	29 Sep 2023 11:58:15	U	SMPL
O4648-01LX5 - 1		48	-0.1506	-	7470A	PPB	29 Sep 2023 12:00:31	U	SMPL
O4648-01A - 1		18865	3.8298	-	7470A	PPB	29 Sep 2023 12:02:47	U	SMPL
O4641-03LX5 - 1		75	-0.1449	-	7470A	PPB	29 Sep 2023 12:05:02	U	SMPL
O4641-03A - 1		19324	3.9269	-	7470A	PPB	29 Sep 2023 12:07:18	U	SMPL
O4645-01 - 1	01	122779	25.8114	-	7470A	PPB	29 Sep 2023 12:14:33	U	SMPL
O4645-01DLX5 - 1		23827	4.8795	-	7470A	PPB	29 Sep 2023 12:20:13	U	SMPL
CCV87 - 1	CCV87	25238	5.1780	-	7470A	PPB	29 Sep 2023 12:22:28	U	SMPL
CCB87 - 1	CCB87	395	-0.0772	-	7470A	PPB	29 Sep 2023 12:24:44	U	SMPL