



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127542

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	09/29/2023	11:00
2	O4641-01		1	11	100.00	>212.0	>212.0	09/29/2023	11:30
3	O4641-02		1	12	100.00	>212.0	>212.0	09/29/2023	12:00
4	O4646-04		1	4	18.00	64.4	64.1	09/29/2023	12:30
5	O4646-04DUP		1	4	18.00	64.4	64.1	09/29/2023	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16127542

WorkList Name : fp s-9-29

WorkList ID : 174080

Department : Wet-Chemistry

Date : 09-29-2023 08:12:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4641-01 A	BS-OIL-DRUM	Solid	Flash Point	Cool 4 deg C	PSEG03	E32	09/27/2023	1010B
O4641-02 A	BS-OIL-TOTE	Solid	Flash Point	Cool 4 deg C	PSEG03	E32	09/27/2023	1010B
O4646-04	22820	Solid	Flash Point	Cool 4 deg C	PSEG03	E11	09/27/2023	1010B

Date/Time 09/29/2023 10:00
 Raw Sample Received by: RM CWS
 Raw Sample Relinquished by: JALCO2C1

Date/Time 09/29/2023 13:10
 Raw Sample Received by: JALCO2C1
 Raw Sample Relinquished by: RM CWS