



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/2/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 09/29/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/30/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127547

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4274-06	RT2683	1	1.14	8.80	9.94	6.5	60.9	GEL MATRIX
O4620-09	STONE-IN	37	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O4620-12	STONE-OUT	38	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O4657-01	U-2 (0-2IN)	2	1.19	8.57	9.76	7.62	75.0	
O4657-02	U-2 (2-6IN)	3	1.12	8.40	9.52	7.95	81.3	
O4657-03	U-1 (0-2IN)	4	1.16	8.77	9.93	7.34	70.5	
O4657-04	U-1 (2-6IN)	5	1.14	8.82	9.96	8.32	81.4	
O4657-05	V-2 (0-2IN)	6	1.19	8.65	9.84	8.17	80.7	
O4657-06	V-2 (2-6IN)	7	1.17	8.50	9.67	8.18	82.5	
O4657-07	V-1 (0-2IN)	8	1.16	8.83	9.99	8.31	81.0	
O4657-08	V-1 (2-6IN)	9	1.15	8.49	9.64	8.06	81.4	
O4657-09	W-2 (0-2IN)	10	1.1	8.74	9.84	8.28	82.2	
O4657-10	W-2 (2-6IN)	11	1.17	8.57	9.74	8.25	82.6	
O4657-11	W-1 (0-2IN)	12	1.16	8.81	9.97	8.07	78.4	
O4657-12	W-1 (2-6IN)	13	1.13	8.84	9.97	8.49	83.3	
O4678-01	B1-10.5-11.5	14	1.17	8.53	9.7	9.25	94.7	
O4678-02	B2-11.5-12.5	15	1.15	8.59	9.74	9.33	95.2	
O4684-01	PIPE-SEGMENTS-1	16	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-02	PIPE-SEGMENTS-2	17	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-03	PIPE-SEGMENTS-3	18	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-04	PIPE-SEGMENTS-4	19	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-05	PIPE-SEGMENTS-5	20	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-06	PIPE-SEGMENTS-6	21	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-07	PIPE-SEGMENTS-7	22	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-08	PIPE-SEGMENTS-8	23	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-09	PIPE-SEGMENTS-9	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-10	PIPE-SEGMENTS-10	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-11	PIPE-SEGMENTS-11	26	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/2/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 09/29/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/30/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127547

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4684-12	PIPE-SEGMENTS-12	27	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-13	PIPE-SEGMENTS-13	28	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-14	PIPE-SEGMENTS-14	29	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-15	PIPE-SEGMENTS-15	30	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-16	PIPE-SEGMENTS-16	31	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-17	PIPE-SEGMENTS-17	32	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-18	PIPE-SEGMENTS-18	33	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-19	PIPE-SEGMENTS-19	34	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4684-20	PIPE-SEGMENTS-20	35	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4686-01	WASTE	36	1.17	8.75	9.92	8.67	85.7	
O4687-01	HR-1-092923	39	1.12	8.66	9.78	7.85	77.7	
O4687-02	HR-1-092923-E2	40	1.15	8.81	9.96	8.25	80.6	
O4687-03	HR-3-092923	41	1.19	8.63	9.82	8.82	88.4	
O4687-04	HR-3-092923-E2	42	1.17	8.60	9.77	8.76	88.3	
O4691-01	NB-08-092923	43	1.13	8.85	9.98	8.48	83.1	
O4691-02	NB-08-092923-E2	44	1.19	8.45	9.64	8.49	86.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$