


LB127548

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.6023e-004

C= -2.8860e-001

Rho= 0.9996698

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.079	-0.079	1311	0.000	1311					0
0.2	0.200	0.200	0.000	3050	0.0 %	3050					0
2.5	2.500	2.610	0.110	18089	0.0 %	18089					4
5.0	5.000	4.932	-0.068	32584	0.0 %	32584					-1
7.5	7.500	7.642	0.142	49494	0.0 %	49494					2
10.0	10.000	9.895	-0.105	63556	0.0 %	63556					-1

LB127548 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Sep 2023 13:53:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	40	1311	-	0.0000	7471B	PPB	29 Sep 2023 13:55:08	S	Std
0.2 - 1	50.2	3050	-	0.2000	7471B	PPB	29 Sep 2023 13:57:25	S	Std
2.5 - 1	52.5	18089	-	2.5000	7471B	PPB	29 Sep 2023 13:59:42	S	Std
5.0 - 1	55	32584	-	5.0000	7471B	PPB	29 Sep 2023 14:01:59	S	Std
7.5 - 1	57.5	49494	-	7.5000	7471B	PPB	29 Sep 2023 14:10:32	S	Std
10.0 - 1	610	63556	-	10.0000	7471B	PPB	29 Sep 2023 14:18:11	S	Std
ICV87 - 1	ICV87	25828	3.8497	-	7471B	PPB	29 Sep 2023 14:24:45	U	SMPL
ICB87 - 1	ICB87	1566	-0.0377	-	7471B	PPB	29 Sep 2023 14:31:29	U	SMPL
CCV87 - 1	CCV87	31918	4.8255	-	7471B	PPB	29 Sep 2023 14:33:44	U	SMPL
CCB87 - 1	CCB87	1272	-0.0848	-	7471B	PPB	29 Sep 2023 14:35:59	U	SMPL
CRA - 1	CRA	3047	0.1996	-	7471B	PPB	29 Sep 2023 14:38:13	U	SMPL
HighStd - 1	HighStd	64368	10.0249	-	7471B	PPB	29 Sep 2023 14:40:28	U	SMPL
ChkStd - 1	ChkStd	48820	7.5337	-	7471B	PPB	29 Sep 2023 14:42:44	U	SMPL
PB155998BL - 1	PBS	2018	0.0347	-	7471B	PPB	29 Sep 2023 14:47:10	U	SMPL
PB155998BS - 1	LCSS	23918	3.5437	-	7471B	PPB	29 Sep 2023 14:49:25	U	SMPL
O4666-01 - 1	VNJ-201	5046	0.5199	-	7471B	PPB	29 Sep 2023 14:51:42	U	SMPL
O4668-01 - 1	RBR200045	848	-0.1527	-	7471B	PPB	29 Sep 2023 14:53:58	U	SMPL
O4668-01DUP - 1	RBR200045DUP	786	-0.1627	-	7471B	PPB	29 Sep 2023 14:56:15	U	SMPL
O4668-01MS - 1	RBR200045MS	26818	4.0084	-	7471B	PPB	29 Sep 2023 14:58:32	U	SMPL
O4668-01MSD - 1	RBR200045MSD	25151	3.7413	-	7471B	PPB	29 Sep 2023 15:00:47	U	SMPL
CCV88 - 1	CCV88	32288	4.8848	-	7471B	PPB	29 Sep 2023 15:03:02	U	SMPL
CCB88 - 1	CCB88	2188	0.0620	-	7471B	PPB	29 Sep 2023 15:05:19	U	SMPL
O4668-03 - 1	RBR200007	1615	-0.0298	-	7471B	PPB	29 Sep 2023 15:07:35	U	SMPL
O4668-05 - 1	RBR251697	2790	0.1584	-	7471B	PPB	29 Sep 2023 15:09:50	U	SMPL
O4668-07 - 1	RBR200027	1201	-0.0962	-	7471B	PPB	29 Sep 2023 15:12:06	U	SMPL
O4668-09 - 1	RBR251354	2213	0.0660	-	7471B	PPB	29 Sep 2023 15:14:22	U	SMPL
O4672-01 - 1	VNJ-208	2303	0.0804	-	7471B	PPB	29 Sep 2023 15:16:39	U	SMPL
O4668-01LX5 - 1		204	-0.2559	-	7471B	PPB	29 Sep 2023 15:18:55	U	SMPL
O4668-01A - 1		27143	4.0604	-	7471B	PPB	29 Sep 2023 15:21:13	U	SMPL
CCV89 - 1	CCV89	32559	4.9282	-	7471B	PPB	29 Sep 2023 15:23:29	U	SMPL
CCB89 - 1	CCB89	1083	-0.1151	-	7471B	PPB	29 Sep 2023 15:32:58	U	SMPL