



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127550

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	09/29/2023	13:40
2	O4639-01		1	12	100.00	>212.0	>212.0	09/29/2023	14:10
3	O4639-01DUP		1	12	100.00	>212.0	>212.0	09/29/2023	14:40
4	O4639-02		1	14	100.00	>212.0	>212.0	09/29/2023	15:10
5	O4641-03		1	12	100.00	>212.0	>212.0	09/29/2023	15:40
6	O4643-02		1	13	100.00	>212.0	>212.0	09/29/2023	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

Lb127550

WorkList Name : FP-W-9/26

WorkList ID : 173984

Department : Wet-Chemistry

Date : 09-26-2023 17:08:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4639-01 C	1035	Water	Flash Point	Cool 4 deg C	PSEG03	E31	09/27/2023	1010B
O4639-02 C	1036	Water	Flash Point	Cool 4 deg C	PSEG03	E31	09/27/2023	1010B
O4641-03 C	BS-RAINWATER	Water	Flash Point	Cool 4 deg C	PSEG03	E32	09/27/2023	1010B
O4643-02 C	217	Water	Flash Point	Cool 4 deg C	PSEG03	E32	09/27/2023	1010B

Date/Time 09/26/2023 13:35
 Raw Sample Received by: PM
 Raw Sample Relinquished by: gfl6007

Date/Time 09/29/2023
 Raw Sample Received by: 1645
 Raw Sample Relinquished by: PM