

LB127551


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5008E



A= 0.0000e+000

B= 2.0117e-004

C= -1.4135e-002

Rho= 0.9995952

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.0	0.000	0.030	0.030	219	0.000	219	0				5
0.2	0.200	0.210	0.010	1112	0.0 %	1112					-4
2.5	2.500	2.406	-0.094	12030	0.0 %	12030					3
5.0	5.000	5.158	0.158	25713	0.0 %	25713					-2
7.5	7.500	7.335	-0.165	36530	0.0 %	36530					1
10.0	10.000	10.062	0.062	50086	0.0 %	50086					

LB127551 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Sep 2023 15:40:54

Sample ID	Extended ID	u Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	219	-	0.000	7470A	PPB	29 Sep 2023 15:41:49	S	Std
0.2 - 1	50.2	1112	-	0.200	7470A	PPB	29 Sep 2023 15:46:46	S	Std
2.5 - 1	52.5	12030	-	2.500	7470A	PPB	29 Sep 2023 15:49:02	S	Std
5.0 - 1	55	25713	-	5.000	7470A	PPB	29 Sep 2023 15:51:19	S	Std
7.5 - 1	57.5	36530	-	7.500	7470A	PPB	29 Sep 2023 15:53:35	S	Std
10.0 - 1	58	50086	-	10.000	7470A	PPB	29 Sep 2023 15:58:25	S	Std
ICV88 - 1	ICV88	18012	3.6093	-	7470A	PPB	29 Sep 2023 16:01:04	U	SMPL
ICB88 - 1	ICB88	59	-0.0023	-	7470A	PPB	29 Sep 2023 16:03:19	U	SMPL
CCV90 - 1	CCV90	25577	5.1311	-	7470A	PPB	29 Sep 2023 16:05:34	U	SMPL
CCB90 - 1	CCB90	65	-0.0011	-	7470A	PPB	29 Sep 2023 16:07:49	U	SMPL
CRA - 1	CRA	1104	0.2080	-	7470A	PPB	29 Sep 2023 16:10:04	U	SMPL
HighStd - 1	HighStd	48864	9.8157	-	7470A	PPB	29 Sep 2023 16:12:18	U	SMPL
ChkStd - 1	ChkStd	33536	6.7322	-	7470A	PPB	29 Sep 2023 16:14:34	U	SMPL
PB155956BL - 1	PBW	-14	-0.0170	-	7470A	PPB	29 Sep 2023 16:16:50	U	SMPL
PB155956BS - 1	LCSW	19555	3.9197	-	7470A	PPB	29 Sep 2023 16:21:36	U	SMPL
O4676-01 - 1	SCF	99	0.0058	-	7470A	PPB	29 Sep 2023 16:23:52	U	SMPL
O4627-02 - 1	280-300FB-BACKFILL-01	392	0.0647	-	7470A	PPB	29 Sep 2023 16:26:08	U	SMPL
O4627-04 - 1	280-300FB-BACKFILL-02	165	0.0191	-	7470A	PPB	29 Sep 2023 16:28:25	U	SMPL
O4627-06 - 1	280-300FB-BACKFILL-03	148	0.0156	-	7470A	PPB	29 Sep 2023 16:30:42	U	SMPL
O4627-08 - 1	280-300FB-BACKFILL-04	139	0.0138	-	7470A	PPB	29 Sep 2023 16:32:57	U	SMPL
CCV91 - 1	CCV91	25618	5.1394	-	7470A	PPB	29 Sep 2023 16:35:12	U	SMPL
CCB91 - 1	CCB91	94	0.0048	-	7470A	PPB	29 Sep 2023 16:37:27	U	SMPL
O4627-10 - 1	280-300FB-BACKFILL-05	115	0.0090	-	7470A	PPB	29 Sep 2023 16:39:42	U	SMPL
O4627-12 - 1	280-300FB-BACKFILL-06	94	0.0048	-	7470A	PPB	29 Sep 2023 16:41:57	U	SMPL
O4627-14 - 1	280-300FB-BACKFILL-07	61	-0.0019	-	7470A	PPB	29 Sep 2023 16:44:13	U	SMPL
O4627-16 - 1	280-300FB-BACKFILL-08	93	0.0046	-	7470A	PPB	29 Sep 2023 16:46:28	U	SMPL
O4627-18 - 1	280-300FB-BACKFILL-09	176	0.0213	-	7470A	PPB	29 Sep 2023 16:48:44	U	SMPL
O4627-20 - 1	280-300FB-BACKFILL-10	165	0.0191	-	7470A	PPB	29 Sep 2023 16:51:00	U	SMPL
O4627-22 - 1	280-300FB-BACKFILL-11	151	0.0162	-	7470A	PPB	29 Sep 2023 16:53:16	U	SMPL
O4627-24 - 1	280-300FB-BACKFILL-12	123	0.0106	-	7470A	PPB	29 Sep 2023 16:55:33	U	SMPL
O4627-26 - 1	280-300FB-BACKFILL-13	127	0.0114	-	7470A	PPB	29 Sep 2023 16:57:51	U	SMPL
O4627-28 - 1	280-300FB-BACKFILL-14	150	0.0160	-	7470A	PPB	29 Sep 2023 17:00:06	U	SMPL
CCV92 - 1	CCV92	25520	5.1197	-	7470A	PPB	29 Sep 2023 17:02:21	U	SMPL
CCB92 - 1	CCB92	111	0.0082	-	7470A	PPB	29 Sep 2023 17:04:36	U	SMPL
O4627-28DUP - 1	280-300FB-BACKFILL-14DUP	256	0.0374	-	7470A	PPB	29 Sep 2023 17:06:52	U	SMPL
O4627-28MS - 1	280-300FB-BACKFILL-14MS	18382	3.6837	-	7470A	PPB	29 Sep 2023 17:09:07	U	SMPL
O4627-28MSD - 1	280-300FB-BACKFILL-14MSD	18954	3.7988	-	7470A	PPB	29 Sep 2023 17:11:23	U	SMPL
O4627-28LX5 - 1		-8	-0.0157	-	7470A	PPB	29 Sep 2023 17:13:38	U	SMPL
O4627-28A - 1		19846	3.9782	-	7470A	PPB	29 Sep 2023 17:15:54	U	SMPL
PB155956TB - 1	PB155948TB	53	-0.0035	-	7470A	PPB	29 Sep 2023 17:18:11	U	SMPL
CCV93 - 1	CCV93	25787	5.1734	-	7470A	PPB	29 Sep 2023 17:20:27	U	SMPL
CCB93 - 1	CCB93	153	0.0166	-	7470A	PPB	29 Sep 2023 17:22:44	U	SMPL