

LB127566


7471B

INSTRUMENT ID: 1

Linear

 μ Abs.:

69627



A= 0.0000e+000

B= 1.4418e-004

C= 8.7874e-003

Rho= 0.9997621

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.042	0.042	227	0.000	227					9.0
0.2	0.200	0.212	0.012	1406	0.0 %	1406					6
2.5	2.500	2.397	-0.103	16565	0.0 %	16565					4
5.0	5.000	5.111	0.111	35386	0.0 %	35386					2
7.5	7.500	7.391	-0.109	51201	0.0 %	51201					1
10.0	10.000	10.048	0.048	69627	0.0 %	69627					0

LB127566 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Oct 2023 13:30:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	227	-	0.0000	7471B	PPB	02 Oct 2023 13:33:21	S	Std
0.2 - 1	512	1406	-	0.2000	7471B	PPB	02 Oct 2023 13:35:37	S	Std
2.5 - 1	525	16565	-	2.5000	7471B	PPB	02 Oct 2023 13:37:54	S	Std
5.0 - 1	55	35386	-	5.0000	7471B	PPB	02 Oct 2023 13:40:11	S	Std
7.5 - 1	575	51201	-	7.5000	7471B	PPB	02 Oct 2023 13:42:27	S	Std
10.0 - 1	518	69627	-	10.0000	7471B	PPB	02 Oct 2023 13:51:17	S	Std
ICV90 - 1	ICV90	27508	3.9750	-	7471B	PPB	02 Oct 2023 13:53:58	U	SMPL
ICB90 - 1	ICB90	191	0.0363	-	7471B	PPB	02 Oct 2023 13:56:12	U	SMPL
CCV97 - 1	CCV97	34972	5.0512	-	7471B	PPB	02 Oct 2023 13:58:27	U	SMPL
CCB97 - 1	CCB97	193	0.0366	-	7471B	PPB	02 Oct 2023 14:00:42	U	SMPL
CRA - 1	CRA	1380	0.2078	-	7471B	PPB	02 Oct 2023 14:02:57	U	SMPL
HighStd - 1	HighStd	70203	10.1309	-	7471B	PPB	02 Oct 2023 14:05:11	U	SMPL
ChkStd - 1	ChkStd	47998	6.9293	-	7471B	PPB	02 Oct 2023 14:07:26	U	SMPL
PB156050BL - 1	PBW	-41	0.0029	-	7471B	PPB	02 Oct 2023 14:09:47	U	SMPL
PB156050BS - 1	LCSS	29597	4.2762	-	7471B	PPB	02 Oct 2023 14:12:05	U	SMPL
O4657-01 - 1	U-2(0-2IN)	27378	3.9562	-	7471B	PPB	02 Oct 2023 14:14:21	U	SMPL
O4657-03 - 1	U-1(0-2IN)	22987	3.3231	-	7471B	PPB	02 Oct 2023 14:16:37	U	SMPL
O4657-05 - 1	V-2(0-2IN)	9813	1.4237	-	7471B	PPB	02 Oct 2023 14:18:54	U	SMPL
O4657-07 - 1	V-1(0-2IN)	13791	1.9972	-	7471B	PPB	02 Oct 2023 14:21:11	U	SMPL
O4657-09 - 1	W-2(0-2IN)	12449	1.8037	-	7471B	PPB	02 Oct 2023 14:23:25	U	SMPL
CCV98 - 1	CCV98	34757	5.0202	-	7471B	PPB	02 Oct 2023 14:25:41	U	SMPL
CCB98 - 1	CCB98	226	0.0414	-	7471B	PPB	02 Oct 2023 14:27:56	U	SMPL
O4657-11 - 1	W-1(0-2IN)	11591	1.6800	-	7471B	PPB	02 Oct 2023 14:30:11	U	SMPL
O4686-01 - 1	WASTE	11201	1.6238	-	7471B	PPB	02 Oct 2023 14:32:26	U	SMPL
O4687-01 - 1	HR-1-092923	2911	0.4285	-	7471B	PPB	02 Oct 2023 14:34:41	U	SMPL
O4687-03 - 1	HR-3-092923	1384	0.2083	-	7471B	PPB	02 Oct 2023 14:41:47	U	SMPL
O4687-03DUP - 1	HR-3-092923DUP	1488	0.2233	-	7471B	PPB	02 Oct 2023 14:44:04	U	SMPL
O4687-03MS - 1	HR-3-092923MS	31082	4.4903	-	7471B	PPB	02 Oct 2023 14:46:19	U	SMPL
O4687-03MSD - 1	HR-3-092923MSD	28731	4.1513	-	7471B	PPB	02 Oct 2023 14:48:35	U	SMPL
O4691-01 - 1	NB-08-092923	2183	0.3235	-	7471B	PPB	02 Oct 2023 14:50:52	U	SMPL
O4687-03LX5 - 1		303	0.0525	-	7471B	PPB	02 Oct 2023 14:53:09	U	SMPL
O4687-03A - 1		29269	4.2289	-	7471B	PPB	02 Oct 2023 14:55:24	U	SMPL
CCV99 - 1	CCV99	34942	5.0468	-	7471B	PPB	02 Oct 2023 14:57:39	U	SMPL
CCB99 - 1	CCB99	219	0.0404	-	7471B	PPB	02 Oct 2023 14:59:55	U	SMPL