



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127567

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/02/2023	08:15
2	CAL2	1	Water	NA	NA	20.6	7.00	10/02/2023	08:16
3	CAL3	1	Water	NA	NA	20.7	10.02	10/02/2023	08:19
4	ICV	1	Water	NA	NA	20.7	7.01	10/02/2023	08:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/02/2023	08:25
6	O4667-01	1	Water	NA	NA	20.1	7.05	10/02/2023	08:30
7	O4682-01	1	Water	NA	NA	20.1	5.94	10/02/2023	08:35
8	O4682-03	1	Water	NA	NA	20.3	6.70	10/02/2023	08:40
9	O4682-05	1	Water	NA	NA	20.5	6.73	10/02/2023	08:45
10	O4682-07	1	Water	NA	NA	20.5	6.73	10/02/2023	08:46
11	O4682-07DUP	1	Water	NA	NA	20.5	6.74	10/02/2023	08:47
12	CCV2	1	Water	NA	NA	20.7	12.02	10/02/2023	08:50

1312756X

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o4667

WorkList ID : 174136

Department : Wet-Chemistry

Date : 10-02-2023 07:49:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4667-01	FRAC-TANK-A2970	Water	pH	Cool 4 deg C	PSEG03	E41	09/28/2023	9040C
O4682-01	MW-1	Water	pH	Cool 4 deg C	LOCK01	F11	09/28/2023	9040C
O4682-03	MW-2	Water	pH	Cool 4 deg C	LOCK01	F11	09/28/2023	9040C
O4682-05	MW-3	Water	pH	Cool 4 deg C	LOCK01	F11	09/28/2023	9040C
O4682-07	MW-4	Water	pH	Cool 4 deg C	LOCK01	F11	09/28/2023	9040C

Date/Time 10/02/23 08:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10/02/23

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

11:00

[Signature]

[Signature]