

LB127573


SFAM01.1

INSTRUMENT ID: CY1

Linear

 μ Abs.:

53514



A= 0.0000e+000

B= 1.8452e-004 slope

C= -1.6278e-002 intercept

Rho= 0.9995888

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.00	0.000	0.013	0.013	161	0.000	161					-7
0.05	0.050					567	0				-5
0.20	0.200	0.186	-0.014	1096	0.0 %	1096					2
2.50	2.500	2.385	-0.115	13015	0.0 %	13015					2
5.00	5.000	5.115	0.115	27810	0.0 %	27810					1
7.50	7.500	7.642	0.142	41505	0.0 %	41505					
10.0	10.000	9.858	-0.142	53514	0.0 %	53514					

LB127573
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	161	-		0 SFAM01.1	PPB	10/2/2023 16:12	Std
	0.2 S01	1096	-		0.2 SFAM01.1	PPB	10/2/2023 16:14	Std
	2.5 S02	13015	-		2.5 SFAM01.1	PPB	10/2/2023 16:17	Std
	5 S03	27810	-		5 SFAM01.1	PPB	10/2/2023 16:19	Std
	7.5 S04	41505	-		7.5 SFAM01.1	PPB	10/2/2023 16:21	Std
	10 S05	53514	-		10 SFAM01.1	PPB	10/2/2023 16:23	Std
ICV091	ICV091	21598	3.969 -		SFAM01.1	PPB	10/2/2023 16:26	SMPL
ICB091	ICB091	98	0.0018 -		SFAM01.1	PPB	10/2/2023 16:28	SMPL
CCV001	CCV001	27688	5.0927 -		SFAM01.1	PPB	10/2/2023 16:31	SMPL
CCB001	CCB001	98	0.0018 -		SFAM01.1	PPB	10/2/2023 16:33	SMPL
PB155954BL	PBW954	94	0.0011 -		SFAM01.1	PPB	10/2/2023 16:35	SMPL
PB155954TB	LEB954	120	0.0059 -		SFAM01.1	PPB	10/2/2023 16:37	SMPL
O4593-01	EZYX9	149	0.0112 -		SFAM01.1	PPB	10/2/2023 16:40	SMPL
O4593-02	EZYY0	166	0.0144 -		SFAM01.1	PPB	10/2/2023 16:42	SMPL
O4593-03	EZYY1	190	0.0188 -		SFAM01.1	PPB	10/2/2023 16:44	SMPL
O4593-04	EZYY2	173	0.0156 -		SFAM01.1	PPB	10/2/2023 16:46	SMPL
O4593-05	EZYY3	128	0.0073 -		SFAM01.1	PPB	10/2/2023 16:49	SMPL
O4593-06	EZYY4	128	0.0073 -		SFAM01.1	PPB	10/2/2023 16:51	SMPL
O4593-07	EZYY4D	177	0.0164 -		SFAM01.1	PPB	10/2/2023 16:53	SMPL
O4593-08	EZYY4S	5446	0.9886 -		SFAM01.1	PPB	10/2/2023 16:55	SMPL
PB155955BL	PBW955	66	-0.0041 -		SFAM01.1	PPB	10/2/2023 16:58	SMPL
PB155955TB	LEB955	124	0.0066 -		SFAM01.1	PPB	10/2/2023 17:00	SMPL
O4611-01	MH1A01	213	0.023 -		SFAM01.1	PPB	10/2/2023 17:02	SMPL
O4611-02	MH1A02	293	0.0378 -		SFAM01.1	PPB	10/2/2023 17:05	SMPL
O4611-03	MH1A03	242	0.0284 -		SFAM01.1	PPB	10/2/2023 17:07	SMPL
O4611-04	MH1A03D	483	0.0728 -		SFAM01.1	PPB	10/2/2023 17:09	SMPL
O4611-05	MH1A03S	6294	1.1451 -		SFAM01.1	PPB	10/2/2023 17:11	SMPL
CCV002	CCV002	27549	5.067 -		SFAM01.1	PPB	10/2/2023 17:14	SMPL
CCB002	CCB002	126	0.007 -		SFAM01.1	PPB	10/2/2023 17:16	SMPL