



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127574

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP104486
Turbidity Calibration std, 20NTU	WP104481
Turbidity Calibration std, 40NTU	WP104480
Turbidity Calibration std, 1NTU	WP104484
Turbidity Calibration std, 5NTU	WP104483
Turbidity Calibration - CCV std, 10 NTU	WP104482
10 NTU Standard 500 ml	W2971

Intercept: 0.1742

Slope: 1.0986

Regression: 0.9994

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.160	-0.01		09/30/2023	08:00
2	CAL2	1	1	1.250	0.98	-2.1	09/30/2023	08:04
3	CAL3	5	1	5.530	4.88	-2.5	09/30/2023	08:08
4	CAL4	10	1	12.000	10.76	7.6	09/30/2023	08:12
5	CAL5	20	1	21.200	19.14	-4.3	09/30/2023	08:16
6	CAL6	40	1	44.400	40.26	0.6	09/30/2023	08:20



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127574

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.500	10.309	09/30/2023	08:24
2	ICB		1	0.164	-0.009	09/30/2023	08:28
3	CCV1	10	1	10.700	9.581	09/30/2023	08:32
4	CCB1		1	0.162	-0.011	09/30/2023	08:36
5	LB127574BL		1	0.183	0.008	09/30/2023	08:40
6	O4349-33		1	18.100	16.317	09/30/2023	08:44
7	O4682-01		1	1.010	0.761	09/30/2023	08:48
8	O4682-01DUP		1	1.020	0.770	09/30/2023	08:52
9	O4682-03		1	12.800	11.493	09/30/2023	08:56
10	O4682-05		1	1.800	1.480	09/30/2023	09:00
11	O4682-07		1	1.330	1.052	09/30/2023	09:04
12	CCV2	10	1	11.400	10.218	09/30/2023	09:08
13	CCB2		1	0.160	-0.013	09/30/2023	09:12

WORKLIST(Hardcopy Internal Chain)

613127579

WorkList Name : TURB-092923

WorkList ID : 174112

Department : Wet-Chemistry

Date : 09-29-2023 12:06:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4682-01	MW-1	Water	Turbidity	Cool 4 deg C	LOCK01	F11	09/28/2023	SM2130 B
O4682-03	MW-2	Water	Turbidity	Cool 4 deg C	LOCK01	F11	09/28/2023	SM2130 B
O4682-05	MW-3	Water	Turbidity	Cool 4 deg C	LOCK01	F11	09/28/2023	SM2130 B
O4682-07	MW-4	Water	Turbidity	Cool 4 deg C	LOCK01	F11	09/28/2023	SM2130 B
O4349-33	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	CHEM02	QA Of	09/28/2023	SM2130 B
O4349-34	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	CHEM02	QA Of	09/05/2023	SM2130 B
							09/05/2023	180.1

613127579

Date/Time 09/29/23 07:45
 Raw Sample Received by: RMA
 Raw Sample Relinquished by: RMA

Date/Time 09/29/23 10:00
 Raw Sample Received by: RMA
 Raw Sample Relinquished by: RMA