



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127575

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP104486
Turbidity Calibration std, 20NTU	WP104481
Turbidity Calibration std, 40NTU	WP104480
Turbidity Calibration std, 1NTU	WP104484
Turbidity Calibration std, 5NTU	WP104483
Turbidity Calibration - CCV std, 10 NTU	WP104482

Intercept: 0.1742

Slope: 1.0986

Regression: 0.9994

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.160	-0.01		09/30/2023	08:00
2	CAL2	1	1	1.250	0.98	-2.1	09/30/2023	08:04
3	CAL3	5	1	5.530	4.88	-2.5	09/30/2023	08:08
4	CAL4	10	1	12.000	10.76	7.6	09/30/2023	08:12
5	CAL5	20	1	21.200	19.14	-4.3	09/30/2023	08:16
6	CAL6	40	1	44.400	40.26	0.6	09/30/2023	08:20



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127575

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.500	10.309	09/30/2023	08:24
2	ICB		1	0.164	-0.009	09/30/2023	08:28
3	CCV1	10	1	10.700	9.581	09/30/2023	08:32
4	CCB1		1	0.162	-0.011	09/30/2023	08:36
5	LB127575BL		1	0.183	0.008	09/30/2023	08:40
6	O4349-34		1	18.100	16.317	09/30/2023	08:44
7	CCV2	10	1	11.400	10.218	09/30/2023	08:48
8	CCB2		1	0.160	-0.013	09/30/2023	08:52