



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127577

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/02/2023	10:30
2	O4667-01		1	12	100.00	>212.0	>212.0	10/02/2023	11:00
3	O4667-01DUP		1	12	100.00	>212.0	>212.0	10/02/2023	11:30
4	O4671-08		1	14	100.00	>212.0	>212.0	10/02/2023	12:00
5	O4671-09		1	13	100.00	>212.0	>212.0	10/02/2023	12:30
6	O4671-10		1	14	100.00	>212.0	>212.0	10/02/2023	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127577

WorkList Name : fp-9-29

WorkList ID : 174117

Department : Wet-Chemistry

Date : 09-29-2023 13:01:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4667-01	FRAC-TANK-A2970	Water	Flash Point	Cool 4 deg C	PSEG03	E41	09/28/2023	1010B
O4671-08	COMP-3	Water	Flash Point	Cool 4 deg C	PSEG03	E61	09/28/2023	1010B
O4671-09	COMP-4	Water	Flash Point	Cool 4 deg C	PSEG03	E61	09/28/2023	1010B
O4671-10	COMP-5	Water	Flash Point	Cool 4 deg C	PSEG03	E61	09/28/2023	1010B

Date/Time

10/02/2023 10:15

Raw Sample Received by:

RM (wcc)

Raw Sample Relinquished by:

Porter

Date/Time

10/02/2023

Raw Sample Received by:

RM

Raw Sample Relinquished by:

RM (wcc)