

LB127583


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40705



A= 0.0000e+000

B= 2.4465e-004

C= 2.1240e-002

Rho= 0.9997947

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.035	0.035	57	0.000	57	0				
0.2	0.200	0.188	-0.012	683	0.0 %	683					-6
2.5	2.500	2.392	-0.108	9690	0.0 %	9690					-4
5.0	5.000	5.138	0.138	20914	0.0 %	20914					3
7.5	7.500	7.467	-0.033	30436	0.0 %	30436					0
10.0	10.000	9.980	-0.020	40705	0.0 %	40705					0

LB127583 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Oct 2023 08:40:20

Sample ID	Extended ID	μ Abs	Conc	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7470A	PPB	03 Oct 2023 08:48:54	S	Std
0.2 - 1	50.2	683	-	0.2000	7470A	PPB	03 Oct 2023 08:51:11	S	Std
2.5 - 1	52.5	9690	-	2.5000	7470A	PPB	03 Oct 2023 08:53:28	S	Std
5.0 - 1	55	20914	-	5.0000	7470A	PPB	03 Oct 2023 08:55:45	S	Std
7.5 - 1	57.5	30438	-	7.5000	7470A	PPB	03 Oct 2023 08:58:01	S	Std
10.0 - 1	510	40705	-	10.0000	7470A	PPB	03 Oct 2023 09:00:17	S	Std
ICV92 - 1	ICV92	15171	3.7328	-	7470A	PPB	03 Oct 2023 09:07:54	U	SMPL
ICB92 - 1	ICB92	-78	0.0022	-	7470A	PPB	03 Oct 2023 09:10:09	U	SMPL
CCV03 - 1	CCV03	20664	5.0766	-	7470A	PPB	03 Oct 2023 09:12:24	U	SMPL
CCB03 - 1	CCB03	-107	-0.0049	-	7470A	PPB	03 Oct 2023 09:14:38	U	SMPL
CRA - 1	CRA	648	0.1798	-	7470A	PPB	03 Oct 2023 09:16:54	U	SMPL
HighStd - 1	HighStd	40162	9.8467	-	7470A	PPB	03 Oct 2023 09:19:09	U	SMPL
ChkStd - 1	ChkStd	28556	7.0074	-	7470A	PPB	03 Oct 2023 09:21:24	U	SMPL
PB156031BL - 1	PBW	-186	-0.0243	-	7470A	PPB	03 Oct 2023 09:23:41	U	SMPL
PB156031BS - 1	LCSW	18042	4.4351	-	7470A	PPB	03 Oct 2023 09:25:58	U	SMPL
O4647-01 - 1	S-1	-55	0.0078	-	7470A	PPB	03 Oct 2023 09:28:14	U	SMPL
O4647-02 - 1	S-2	13	0.0244	-	7470A	PPB	03 Oct 2023 09:30:30	U	SMPL
O4647-03 - 1	S-3	20	0.0261	-	7470A	PPB	03 Oct 2023 09:32:47	U	SMPL
O4647-04 - 1	S-4	6	0.0227	-	7470A	PPB	03 Oct 2023 09:35:04	U	SMPL
O4647-05 - 1	S-5	28	0.0281	-	7470A	PPB	03 Oct 2023 09:37:19	U	SMPL
CCV04 - 1	CCV04	20974	5.1524	-	7470A	PPB	03 Oct 2023 09:39:34	U	SMPL
CCB04 - 1	CCB04	-79	0.0019	-	7470A	PPB	03 Oct 2023 09:41:49	U	SMPL
O4647-06 - 1	S-6	48	0.0330	-	7470A	PPB	03 Oct 2023 09:44:04	U	SMPL
O4647-07 - 1	S-7	10	0.0237	-	7470A	PPB	03 Oct 2023 09:46:19	U	SMPL
O4647-08 - 1	S-8	42	0.0315	-	7470A	PPB	03 Oct 2023 09:48:34	U	SMPL
O4651-01 - 1	SP1-COMP	33	0.0293	-	7470A	PPB	03 Oct 2023 09:50:49	U	SMPL
O4652-01 - 1	IS1-COMP	21	0.0264	-	7470A	PPB	03 Oct 2023 09:53:05	U	SMPL
O4652-04 - 1	IS2-COMP	39	0.0308	-	7470A	PPB	03 Oct 2023 09:55:21	U	SMPL
O4652-07 - 1	IS3-COMP	25	0.0274	-	7470A	PPB	03 Oct 2023 09:57:37	U	SMPL
O4652-10 - 1	IS4-COMP	19	0.0259	-	7470A	PPB	03 Oct 2023 09:59:54	U	SMPL
O4652-13 - 1	IS5-COMP	30	0.0286	-	7470A	PPB	03 Oct 2023 10:02:10	U	SMPL
O4652-16 - 1	IS6-COMP	37	0.0303	-	7470A	PPB	03 Oct 2023 10:04:26	U	SMPL
CCV05 - 1	CCV05	21076	5.1774	-	7470A	PPB	03 Oct 2023 10:06:42	U	SMPL
CCB05 - 1	CCB05	-58	0.0071	-	7470A	PPB	03 Oct 2023 10:08:57	U	SMPL
O4660-01 - 1	BLDG-39-ROOM-5-75-092723JE18-TCLP	10375	2.5594	-	7470A	PPB	03 Oct 2023 10:11:12	U	SMPL
O4666-03 - 1	VNJ-201	-2	0.0208	-	7470A	PPB	03 Oct 2023 10:13:28	U	SMPL
O4666-03DUP - 1	VNJ-201DUP	45	0.0322	-	7470A	PPB	03 Oct 2023 10:15:43	U	SMPL
O4666-03MS - 1	VNJ-201MS	16861	4.1462	-	7470A	PPB	03 Oct 2023 10:18:00	U	SMPL
O4666-03MSD - 1	VNJ-201MSD	16180	3.9796	-	7470A	PPB	03 Oct 2023 10:20:16	U	SMPL
PB156075BL - 1	PBW	-31	0.0137	-	7470A	PPB	03 Oct 2023 10:22:32	U	SMPL
PB156075BS - 1	LCSW	15755	3.8756	-	7470A	PPB	03 Oct 2023 10:24:48	U	SMPL
O4671-01 - 1	COMP-1	-46	0.0100	-	7470A	PPB	03 Oct 2023 10:27:05	U	SMPL
O4671-01DUP - 1	COMP-1DUP	22	0.0266	-	7470A	PPB	03 Oct 2023 10:29:22	U	SMPL
O4671-01MS - 1	COMP-1MS	15596	3.8367	-	7470A	PPB	03 Oct 2023 10:31:37	U	SMPL
CCV06 - 1	CCV06	21024	5.1647	-	7470A	PPB	03 Oct 2023 10:33:53	U	SMPL
CCB06 - 1	CCB06	-83	0.0009	-	7470A	PPB	03 Oct 2023 10:36:08	U	SMPL
O4671-01MSD - 1	COMP-1MSD	16457	4.0474	-	7470A	PPB	03 Oct 2023 10:38:23	U	SMPL
O4671-02 - 1	COMP-2	-17	0.0171	-	7470A	PPB	03 Oct 2023 10:40:39	U	SMPL
O4671-06 - 1	8450	18	0.0256	-	7470A	PPB	03 Oct 2023 10:42:54	U	SMPL
PB156076BL - 1	PBW	27	0.0278	-	7470A	PPB	03 Oct 2023 10:45:09	U	SMPL
PB156076BS - 1	LCSW	15963	3.9265	-	7470A	PPB	03 Oct 2023 10:47:25	U	SMPL
O4671-08 - 1	COMP-3	69	0.0381	-	7470A	PPB	03 Oct 2023 10:49:41	U	SMPL
O4671-09 - 1	COMP-4	52	0.0340	-	7470A	PPB	03 Oct 2023 10:51:57	U	SMPL
O4671-10 - 1	COMP-5	67	0.0376	-	7470A	PPB	03 Oct 2023 10:54:14	U	SMPL
O4671-10DUP - 1	COMP-5DUP	78	0.0403	-	7470A	PPB	03 Oct 2023 10:56:30	U	SMPL
O4671-10MS - 1	COMP-5MS	15016	3.6948	-	7470A	PPB	03 Oct 2023 10:58:45	U	SMPL
CCV07 - 1	CCV07	19863	4.8806	-	7470A	PPB	03 Oct 2023 11:01:01	U	SMPL
CCB07 - 1	CCB07	-90	-0.0008	-	7470A	PPB	03 Oct 2023 11:03:16	U	SMPL
O4671-10MSD - 1	COMP-5MSD	15700	3.8622	-	7470A	PPB	03 Oct 2023 11:05:32	U	SMPL
PB155983TB - 1	PB155983TB	-58	0.0071	-	7470A	PPB	03 Oct 2023 11:07:47	U	SMPL
PB156014TB - 1	PB156014TB	6	0.0227	-	7470A	PPB	03 Oct 2023 11:10:03	U	SMPL
PB155991TB - 1	PB155991TB	16	0.0252	-	7470A	PPB	03 Oct 2023 11:12:18	U	SMPL
O4666-03LX5 - 1		8	0.0232	-	7470A	PPB	03 Oct 2023 11:14:33	U	SMPL
O4666-03A - 1		16333	4.0170	-	7470A	PPB	03 Oct 2023 11:16:50	U	SMPL
O4671-01LX5 - 1		-35	0.0127	-	7470A	PPB	03 Oct 2023 11:19:06	U	SMPL
O4671-01A - 1		17033	4.1883	-	7470A	PPB	03 Oct 2023 11:21:22	U	SMPL
O4671-10LX5 - 1		-14	0.0178	-	7470A	PPB	03 Oct 2023 11:23:38	U	SMPL
O4671-10A - 1		18544	4.5580	-	7470A	PPB	03 Oct 2023 11:25:54	U	SMPL
CCV08 - 1	CCV08	20195	4.9619	-	7470A	PPB	03 Oct 2023 11:28:12	U	SMPL
CCB08 - 1	CCB08	-93	-0.0015	-	7470A	PPB	03 Oct 2023 11:30:28	U	SMPL