


LB127601

7470A

INSTRUMENT ID: CV1

 μ Abs.:

4809E



Max. Conc: —

10.000 PPB

Linear

A= 0.0000e+000

B= 2.0859e-004

C= -9.8652e-002

Rho= 0.9998922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0 D
0.0	0.000	-0.053	-0.053	221	0.000	221					
0.2	0.200	0.200	0.000	1434	0.0 %	1434	0				
2.5	2.500	2.559	0.059	12742	0.0 %	12742					
5.0	5.000	4.979	-0.021	24341	0.0 %	24341					
7.5	7.500	7.581	0.081	36815	0.0 %	36815					
10.0	10.000	9.934	-0.066	48096	0.0 %	48096					

1/0 D
 2
 0
 1
 1

LB127601 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Oct 2023 17:36:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	221	-	0.0000	7470A	PPB	04 Oct 2023 08:03:02	S	Std
0.2 - 1	50.2	1434	-	0.2000	7470A	PPB	04 Oct 2023 08:05:20	S	Std
2.5 - 1	52.5	12742	-	2.5000	7470A	PPB	04 Oct 2023 08:07:37	S	Std
5.0 - 1	55	24341	-	5.0000	7470A	PPB	04 Oct 2023 08:09:53	S	Std
7.5 - 1	57.5	36815	-	7.5000	7470A	PPB	04 Oct 2023 08:12:09	S	Std
10.0 - 1	516	48096	-	10.0000	7470A	PPB	04 Oct 2023 08:16:57	S	Std
ICV93 - 1	ICV93	19640	3.9981	-	7470A	PPB	04 Oct 2023 08:19:42	U	SMPL
ICB93 - 1	ICB93	113	-0.0751	-	7470A	PPB	04 Oct 2023 08:21:56	U	SMPL
CCV09 - 1	CCV09	23355	4.7730	-	7470A	PPB	04 Oct 2023 08:24:11	U	SMPL
CCB09 - 1	CCB09	127	-0.0722	-	7470A	PPB	04 Oct 2023 08:26:26	U	SMPL
CRA - 1	CRA	1428	0.1992	-	7470A	PPB	04 Oct 2023 08:28:41	U	SMPL
HighStd - 1	HighStd	47617	9.8338	-	7470A	PPB	04 Oct 2023 08:30:56	U	SMPL
ChkStd - 1	ChkStd	35505	7.3073	-	7470A	PPB	04 Oct 2023 08:33:10	U	SMPL
PB156083BL - 1	PBW	681	0.0434	-	7470A	PPB	04 Oct 2023 08:35:26	U	SMPL
PB156083BS - 1	LCSW	22083	4.5076	-	7470A	PPB	04 Oct 2023 08:37:42	U	SMPL
O4703-02 - 1	E9066-OB-001	691	0.0455	-	7470A	PPB	04 Oct 2023 08:39:57	U	SMPL
O4703-04 - 1	E9066-OB-002	776	0.0632	-	7470A	PPB	04 Oct 2023 08:42:14	U	SMPL
O4703-06 - 1	E9066-OB-003	833	0.0751	-	7470A	PPB	04 Oct 2023 08:44:31	U	SMPL
O4703-06DUP - 1	E9066-OB-003DUP	872	0.0832	-	7470A	PPB	04 Oct 2023 08:46:48	U	SMPL
O4703-06MS - 1	E9066-OB-003MS	20664	4.2116	-	7470A	PPB	04 Oct 2023 08:49:03	U	SMPL
CCV10 - 1	CCV10	23112	4.7223	-	7470A	PPB	04 Oct 2023 08:51:18	U	SMPL
CCB10 - 1	CCB10	127	-0.0722	-	7470A	PPB	04 Oct 2023 08:53:33	U	SMPL
O4703-06MSD - 1	E9066-OB-003MSD	21005	4.2828	-	7470A	PPB	04 Oct 2023 08:55:48	U	SMPL
PB156092BL - 1	PBW	867	0.0822	-	7470A	PPB	04 Oct 2023 08:58:03	U	SMPL
PB156092BS - 1	LCSW	20385	4.1534	-	7470A	PPB	04 Oct 2023 09:00:19	U	SMPL
O4656-01 - 1	MW-105D	674	0.0419	-	7470A	PPB	04 Oct 2023 09:02:34	U	SMPL
O4656-02 - 1	MW-105I	773	0.0626	-	7470A	PPB	04 Oct 2023 09:04:50	U	SMPL
O4656-03 - 1	MW-106D	752	0.0582	-	7470A	PPB	04 Oct 2023 09:07:06	U	SMPL
O4656-04 - 1	MW-106I	863	0.0814	-	7470A	PPB	04 Oct 2023 09:09:22	U	SMPL
O4656-05 - 1	MW-106S	827	0.0739	-	7470A	PPB	04 Oct 2023 09:11:39	U	SMPL
O4656-05DUP - 1	MW-106SDUP	846	0.0778	-	7470A	PPB	04 Oct 2023 09:13:56	U	SMPL
O4656-06 - 1	O4656-05MS	21407	4.3666	-	7470A	PPB	04 Oct 2023 09:16:12	U	SMPL
CCV11 - 1	CCV11	23379	4.7780	-	7470A	PPB	04 Oct 2023 09:18:27	U	SMPL
CCB11 - 1	CCB11	120	-0.0736	-	7470A	PPB	04 Oct 2023 09:20:43	U	SMPL
O4656-07 - 1	O4656-05MSD	19832	4.0381	-	7470A	PPB	04 Oct 2023 09:22:58	U	SMPL
O4656-09 - 1	MW-103D	179	-0.0613	-	7470A	PPB	04 Oct 2023 09:25:13	U	SMPL
O4656-10 - 1	MW-103I	213	-0.0542	-	7470A	PPB	04 Oct 2023 09:27:28	U	SMPL
O4656-11 - 1	MW-103S	297	-0.0367	-	7470A	PPB	04 Oct 2023 09:29:44	U	SMPL
O4656-12 - 1	MW-104D	256	-0.0453	-	7470A	PPB	04 Oct 2023 09:32:00	U	SMPL
O4656-13 - 1	MW-104I	1088	0.1283	-	7470A	PPB	04 Oct 2023 09:34:16	U	SMPL
O4656-15 - 1	MW-105S	205	-0.0559	-	7470A	PPB	04 Oct 2023 09:36:32	U	SMPL
O4656-16 - 1	MW-102D	932	0.0958	-	7470A	PPB	04 Oct 2023 09:38:49	U	SMPL
O4656-18 - 1	MW-101S	239	-0.0488	-	7470A	PPB	04 Oct 2023 09:41:05	U	SMPL
O4656-19 - 1	MW-101I	246	-0.0473	-	7470A	PPB	04 Oct 2023 09:43:21	U	SMPL
CCV12 - 1	CCV12	24764	5.0662	-	7470A	PPB	04 Oct 2023 09:45:36	U	SMPL
CCB12 - 1	CCB12	123	-0.0730	-	7470A	PPB	04 Oct 2023 09:47:51	U	SMPL
O4656-20 - 1	MW-101D	1545	0.2236	-	7470A	PPB	04 Oct 2023 09:50:07	U	SMPL
O4656-21 - 1	MW-102S	805	0.0693	-	7470A	PPB	04 Oct 2023 09:52:22	U	SMPL
O4656-22 - 1	MW-102I	364	-0.0227	-	7470A	PPB	04 Oct 2023 09:54:38	U	SMPL
O4656-23 - 1	MW-107(20)	296	-0.0369	-	7470A	PPB	04 Oct 2023 09:56:53	U	SMPL
O4656-24 - 1	FB-10022023	190	-0.0590	-	7470A	PPB	04 Oct 2023 09:59:09	U	SMPL
O4685-13 - 1	FIELD-BLANK	889	0.0868	-	7470A	PPB	04 Oct 2023 10:01:25	U	SMPL
CCV13 - 1	CCV13	24764	5.0669	-	7470A	PPB	04 Oct 2023 10:03:40	U	SMPL
CCB13 - 1	CCB13	121	-0.0734	-	7470A	PPB	04 Oct 2023 10:05:56	U	SMPL
PB156065TB - 1	PB156065TB	724	0.0524	-	7470A	PPB	04 Oct 2023 10:08:11	U	SMPL
O4703-06LX5 - 1		248	-0.0469	-	7470A	PPB	04 Oct 2023 10:10:27	U	SMPL
O4703-06A - 1		19876	4.0473	-	7470A	PPB	04 Oct 2023 10:12:42	U	SMPL
O4656-05LX5 - 1		168	-0.0636	-	7470A	PPB	04 Oct 2023 10:14:58	U	SMPL
O4656-05A - 1		20463	4.1697	-	7470A	PPB	04 Oct 2023 10:17:13	U	SMPL
CCV14 - 1	CCV14	24641	5.0412	-	7470A	PPB	04 Oct 2023 10:19:29	U	SMPL
CCB14 - 1	CCB14	111	-0.0755	-	7470A	PPB	04 Oct 2023 10:21:44	U	SMPL
							04 Oct 2023 10:26:57	U	SMPL