



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127611

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/04/2023	10:30
2	O4718-01		1	11	100.00	>212.0	>212.0	10/04/2023	11:00
3	O4718-01DUP		1	12	100.00	>212.0	>212.0	10/04/2023	11:30
4	O4720-03		1	14	100.00	>212.0	>212.0	10/04/2023	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127611

WorkList Name : fp w 10-4 WorkList ID : 174225 Department : Wet-Chemistry Date : 10-04-2023 08:15:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4718-01 F	A455	Water	Flash Point	Cool 4 deg C	PSEG03	D11	10/03/2023	1010B
O4720-03 D	ORA-1041	Water	Flash Point	Cool 4 deg C	PSEG03	E32	10/03/2023	1010B

Date/Time 10/04/2023 10:20
 Raw Sample Received by: RPA cwc
 Raw Sample Relinquished by: JH LWC

Date/Time 10/04/2023 12:40
 Raw Sample Received by: RPA cwc
 Raw Sample Relinquished by: RPA cwc