



Analytical Summary Report

Analysis Method: SM2120 B
Parameter: Color
Run Number: LB127614

Reviewed By: Iwona
Supervisor Review By: Sohil
pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP104037
Calibration std, 5 color unit	WP104038
Calibration std, 10 color unit	WP104039
Calibration std, 15 color unit	WP104040
Calibration std, 20 color unit	WP104041
Calibration std, 25 color unit	WP104042
Calibration std, 30 color unit	WP104043
Calibration std, 40 color unit	WP104044
Calibration std, 50 color unit	WP104045
Calibration std, 60 color unit	WP104046
Calibration std, 70 color unit	WP104047
Calibration std, 100 color unit	WP104048

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB127614BL	LB127614BL	1	5.02	N/A	0.00	10/04/2023	13:10
2	O4349-27	PT-COL-WP	1	1.35	N/A	25.00	10/04/2023	13:15
3	O4715-09	7CS	10	7.98	N/A	50.00	10/04/2023	13:20
4	O4715-09DUP	7CSDUP	10	7.97	N/A	50.00	10/04/2023	13:25

WORKLIST(Hardcopy Internal Chain)

LB 127614

WorkList Name : COLOR-100423

WorkList ID : 174235

Department : Wet-Chemistry

Date : 10-04-2023 10:40:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4349-27	PT-COL-WP	Water	Color	Cool 4 deg C	CHEM02	QA Of	09/05/2023	SM2120 B
O4715-09	7CS	Water	Color	Cool 4 deg C	M&MM01	E41	10/03/2023	SM2120 B

Date/Time 10/04/23 12:40
 Raw Sample Received by: 12/23/23
 Raw Sample Relinquished by: 12/23/23

Date/Time 10/04/23 13:40
 Raw Sample Received by: 12/23/23
 Raw Sample Relinquished by: 12/23/23