



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127615

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	10/04/2023	13:00
2	O4720-01		1	12	100.00	>212.0	>212.0	10/04/2023	13:30
3	O4720-01DUP		1	12	100.00	>212.0	>212.0	10/04/2023	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127615

WorkList Name : f p s 10-4 WorkList ID : 174226 Department : Wet-Chemistry Date : 10-04-2023 08:15:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4720-01	ORA-1042	Solid	Flash Point	Cool 4 deg C	PSEG03	E32	10/03/2023	1010B

Date/Time 10/04/2023 17:45
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 10/04/2023 10:40
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC