


LB127624

SFAM01.1

INSTRUMENT ID: EVJ



Linear

A= 0.0000e+000

B= 1.6757e-004 slope

C= -8.6521e-002 y-intercept

Rho= 0.9993333

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	-0.068	-0.068	109	0.000	109					0/01)
0.05	0.050					364					11
0.20	0.200	0.168	-0.032	1520	0.0 %	1520	0				-16
2.50	2.500	2.521	0.021	15562	0.0 %	15562					1
5.00	5.000	5.272	0.272	31980	0.0 %	31980					5
7.50	7.500	7.333	-0.167	44277	0.0 %	44277					-2
10.0	10.000	9.974	-0.026	60035	0.0 %	60035					0

LB127624
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	109	-		0 SFAM01.1	PPB	10/5/2023 8:03	Std
	0.2 S01	1520	-		0.2 SFAM01.1	PPB	10/5/2023 8:05	Std
	2.5 S02	15562	-		2.5 SFAM01.1	PPB	10/5/2023 8:08	Std
	5 S03	31980	-		5 SFAM01.1	PPB	10/5/2023 8:10	Std
	7.5 S04	44277	-		7.5 SFAM01.1	PPB	10/5/2023 8:12	Std
	10 S05	60035	-		10 SFAM01.1	PPB	10/5/2023 8:18	Std
ICV096	ICV096	23436	3.8406 -		SFAM01.1	PPB	10/5/2023 8:21	SMPL
ICB096	ICB096	99	-0.0699 -		SFAM01.1	PPB	10/5/2023 8:23	SMPL
CCV024	CCV024	32089	5.2906 -		SFAM01.1	PPB	10/5/2023 8:25	SMPL
CCB024	CCB024	41	-0.0797 -		SFAM01.1	PPB	10/5/2023 8:28	SMPL
PB156127BL	PBS127	155	-0.0605 -		SFAM01.1	PPB	10/5/2023 8:30	SMPL
O4482-01	MEZYK8	1982	0.2456 -		SFAM01.1	PPB	10/5/2023 8:32	SMPL
O4482-02	MEZYL0	500	-0.0027 -		SFAM01.1	PPB	10/5/2023 8:34	SMPL
O4482-03	MEZYK9	134	-0.0641 -		SFAM01.1	PPB	10/5/2023 8:37	SMPL
O4482-04	MEZYYW9	184	-0.0557 -		SFAM01.1	PPB	10/5/2023 8:39	SMPL
O4482-05	MEZYM3	4079	0.597 -		SFAM01.1	PPB	10/5/2023 8:41	SMPL
O4482-06	MEZYM4	2246	0.2898 -		SFAM01.1	PPB	10/5/2023 8:43	SMPL
O4482-07	MEZYM5	49287	8.1725 -		SFAM01.1	PPB	10/5/2023 8:46	SMPL
O4482-08	MEZYM6	147	-0.0619 -		SFAM01.1	PPB	10/5/2023 8:48	SMPL
O4482-09	MEZYM6D	199	-0.0532 -		SFAM01.1	PPB	10/5/2023 8:50	SMPL
O4482-10	MEZYM6S	15391	2.4925 -		SFAM01.1	PPB	10/5/2023 8:55	SMPL
O4482-11	MEZYM7	114	-0.0674 -		SFAM01.1	PPB	10/5/2023 8:57	SMPL
O4482-12	MEZYM8	4861	0.728 -		SFAM01.1	PPB	10/5/2023 8:59	SMPL
PB156128BL	PBS128	91	-0.0713 -		SFAM01.1	PPB	10/5/2023 9:02	SMPL
O4529-01	MEZYG8	339	-0.0297 -		SFAM01.1	PPB	10/5/2023 9:04	SMPL
O4529-02	MEZYG9	2441	0.3225 -		SFAM01.1	PPB	10/5/2023 9:06	SMPL
O4529-03	MEZYH0	85914	14.3101 -		SFAM01.1	PPB	10/5/2023 9:08	SMPL
O4529-04	MEZYK5	1011	0.0829 -		SFAM01.1	PPB	10/5/2023 9:11	SMPL
O4529-05	MEZYK5D	1020	0.0844 -		SFAM01.1	PPB	10/5/2023 9:13	SMPL
O4529-06	MEZYK5S	18533	3.0191 -		SFAM01.1	PPB	10/5/2023 9:18	SMPL
CCV025	CCV025	31514	5.1943 -		SFAM01.1	PPB	10/5/2023 9:20	SMPL
CCB025	CCB025	44	-0.0791 -		SFAM01.1	PPB	10/5/2023 9:23	SMPL
O4529-07	MEZYK6	637	0.0202 -		SFAM01.1	PPB	10/5/2023 9:25	SMPL
O4529-08	MEZYK7	7895	1.2364 -		SFAM01.1	PPB	10/5/2023 9:27	SMPL
O4529-09	MEZYY7	4649	0.6925 -		SFAM01.1	PPB	10/5/2023 9:29	SMPL
O4529-10	MEZYY8	3130	0.438 -		SFAM01.1	PPB	10/5/2023 9:32	SMPL
O4529-11	MEZYY5	1920	0.2352 -		SFAM01.1	PPB	10/5/2023 9:34	SMPL
O4529-12	MEZYY6	891	0.0628 -		SFAM01.1	PPB	10/5/2023 9:36	SMPL
O4529-13	MEZYY9	696	0.0301 -		SFAM01.1	PPB	10/5/2023 9:38	SMPL
O4529-03DLX5	MEZYH0	18497	3.013 -		SFAM01.1	PPB	10/5/2023 9:41	SMPL
CCV026	CCV026	31329	5.1633 -		SFAM01.1	PPB	10/5/2023 9:43	SMPL
CCB026	CCB026	43	-0.0793 -		SFAM01.1	PPB	10/5/2023 9:45	SMPL