

LB127627

Test results

Aguakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

10/5/2023 11:48

Reviewed by : 12 Instrument ID : Konelab

Test: Total CN

Sample Id	^{W/L} Result	Dil. 1 +	Response	Errors
ICV1	92.360	0.0	0.071	
ICB1	1.398	0.0	0.001	
CCV1	237.636	0.0	0.184	
CCB1	1.268	0.0	0.001	
PB156155BL	1.011	0.0	0.001	
O4717-01	1.412	0.0	0.001	
O4717-01DUP	1.229	0.0	0.001	
CCV2	239.354	0.0	0.185	
CCB2	1.350	0.0	0.001	

N	9
Mean	64.113
SD	103.2636
CV%	161.06

Aquakem v. 7.2AQ1

Results from time period:

Thu Oct 05 11:34:10 2023

Thu Oct 05 11:41:55 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	1.6995	µg/l	10/5/2023 9:15:40	
5.OPPBCN	A	Total CN	P	5.9663	µg/l	10/5/2023 9:15:41	
10PPBCN	A	Total CN	P	10.2544	µg/l	10/5/2023 9:15:42	
50PPBCN	A	Total CN	P	49.1991	µg/l	10/5/2023 9:15:43	
100PPBCN	A	Total CN	P	98.0982	µg/l	10/5/2023 9:15:44	
250PPBCN	A	Total CN	P	248.6736	µg/l	10/5/2023 9:15:45	
500PPBCN	A	Total CN	P	501.1089	µg/l	10/5/2023 9:15:46	
ICV1	S	Total CN	P	92.3603	µg/l	10/5/2023 11:34:11	
ICB1	S	Total CN	P	1.3979	µg/l	10/5/2023 11:34:12	
CCV1	S	Total CN	P	237.6358	µg/l	10/5/2023 11:34:15	
CCB1	S	Total CN	P	1.2684	µg/l	10/5/2023 11:34:17	
PB156155BL	S	Total CN	P	1.0109	µg/l	10/5/2023 11:34:19	
O4717-01	S	Total CN	P	1.4125	µg/l	10/5/2023 11:34:20	
O4717-01DUP	S	Total CN	P	1.2285	µg/l	10/5/2023 11:41:46	
CCV2	S	Total CN	P	239.3541	µg/l	10/5/2023 11:41:50	
CCB2	S	Total CN	P	1.3498	µg/l	10/5/2023 11:41:53	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

10/5/2023 9:18

Reviewed by : 12

Instrument ID : Konelab

Test Total CN


Accepted 10/5/2023 9:18

Factor 1291

Bias 0

Coeff. of det. 0.999946

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	1.6995	0.0000	-
2	5.0PPBCN	0.004	5.9663	5.0000	18.3
3	10PPBCN	0.008	10.2544	10.0000	2.5
4	50PPBCN	0.038	49.1991	50.0000	-1.6
5	100PPBCN	0.076	98.0982	100.0000	-1.9
6	250PPBCN	0.192	248.6736	250.0000	-0.5
7	500PPBCN	0.388	501.1089	500.0000	0.2

12
10/5/23