



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127633

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/05/2023	11:50
2	CAL2	1	Water	NA	NA	20.7	7.00	10/05/2023	11:51
3	CAL3	1	Water	NA	NA	20.7	10.02	10/05/2023	11:52
4	ICV	1	Water	NA	NA	20.6	7.00	10/05/2023	12:00
5	CCV1	1	Water	NA	NA	20.6	2.01	10/05/2023	12:05
6	O4733-01	1	Solid	20.02	20	21.0	8.33	10/05/2023	12:10
7	O4733-01DUP	1	Solid	20.03	20	21.2	8.34	10/05/2023	12:11
8	O4769-01	1	Solid	20.02	20	20.9	8.40	10/05/2023	12:15
9	CCV2	1	Water	NA	NA	20.6	12.02	10/05/2023	12:20

WB12F633

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s 04733

WorkList ID : 174289

Department : Wet-Chemistry

Date : 10-05-2023 11:40:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4733-01 <i>OK</i>	20231004-WC1	Solid	pH	Cool 4 deg C	WALS01	E41	10/04/2023	9045D
O4769-01 <i>OK</i>	AU-06-10-4-2023	Solid	pH	Cool 4 deg C	PSEG03	E11	10/04/2023	9045D

11:45

Date/Time 10/05/23
Raw Sample Received by: *SP (WC)*
Raw Sample Relinquished by: *SP (WC)*

Date/Time 10/05/23
Raw Sample Received by: *SP (WC)*
Raw Sample Relinquished by: *SP (WC)*