

LB127634

7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

68895

A= 0.0000e+000

B= 1.4447e-004

C= -4.2087e-003

Rho= 0.9998395

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	40D
0.0	0.000	0.019	0.019	159	0.000	159					-
0.2	0.200	0.215	0.015	1518	0.0 %	1518					7
2.5	2.500	2.469	-0.031	17118	0.0 %	17118					-1
5.0	5.000	4.922	-0.078	34095	0.0 %	34095					-2
7.5	7.500	7.626	0.126	52816	0.0 %	52816					2
10.0	10.000	9.949	-0.051	68895	0.0 %	68895					-1

LB127634 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Oct 2023 14:27:24

Sample ID	Extended ID	Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	159	-	0.0000	7471B	PPB	05 Oct 2023 14:46:41	S	Std
0.2 - 1	50.2	1518	-	0.2000	7471B	PPB	05 Oct 2023 14:48:59	S	Std
2.5 - 1	52.5	17118	-	2.5000	7471B	PPB	05 Oct 2023 14:51:16	S	Std
5.0 - 1	55	34095	-	5.0000	7471B	PPB	05 Oct 2023 14:53:33	S	Std
7.5 - 1	57.5	52816	-	7.5000	7471B	PPB	05 Oct 2023 14:55:49	S	Std
10.0 - 1	510	68895	-	10.0000	7471B	PPB	05 Oct 2023 15:04:40	S	Std
ICV97 - 1	ICV97	27160	3.9197	-	7471B	PPB	05 Oct 2023 15:07:25	U	SMPL
ICB97 - 1	ICB97	37	0.0011	-	7471B	PPB	05 Oct 2023 15:09:39	U	SMPL
CCV27 - 1	CCV27	35445	5.1167	-	7471B	PPB	05 Oct 2023 15:11:54	U	SMPL
CCB27 - 1	CCB27	43	0.0020	-	7471B	PPB	05 Oct 2023 15:14:09	U	SMPL
CRA - 1	CRA	1540	0.2183	-	7471B	PPB	05 Oct 2023 15:16:25	U	SMPL
HighStd - 1	HighStd	68819	9.9383	-	7471B	PPB	05 Oct 2023 15:18:39	U	SMPL
ChkStd - 1	ChkStd	51126	7.3822	-	7471B	PPB	05 Oct 2023 15:20:54	U	SMPL
PB156159BL - 1	PBS	-78	-0.0155	-	7471B	PPB	05 Oct 2023 15:23:11	U	SMPL
PB156159BS - 1	LCSS	27996	4.0405	-	7471B	PPB	05 Oct 2023 15:25:28	U	SMPL
O4685-01 - 1	PP-20230928-B4A-0-7.5	4159	0.5967	-	7471B	PPB	05 Oct 2023 15:27:44	U	SMPL
O4685-03 - 1	PP-20230928-B4A-7.5-15	842	0.1174	-	7471B	PPB	05 Oct 2023 15:30:00	U	SMPL
O4685-05 - 1	PP-20230928-B6A-0-7.5	7888	1.1354	-	7471B	PPB	05 Oct 2023 15:32:17	U	SMPL
O4685-07 - 1	PP-20230928-B6A-7.5-15	944	0.1322	-	7471B	PPB	05 Oct 2023 15:34:34	U	SMPL
O4685-09 - 1	PP-20230928-B13-0-7.5	3528	0.5055	-	7471B	PPB	05 Oct 2023 15:36:49	U	SMPL
CCV28 - 1	CCV28	35926	5.1862	-	7471B	PPB	05 Oct 2023 15:39:04	U	SMPL
CCB28 - 1	CCB28	69	0.0058	-	7471B	PPB	05 Oct 2023 15:41:19	U	SMPL
O4685-11 - 1	PP-20230928-B13-7.5-15	4168	0.5980	-	7471B	PPB	05 Oct 2023 15:43:35	U	SMPL
O4693-01 - 1	F-1(0-5)	8522	1.2270	-	7471B	PPB	05 Oct 2023 15:45:50	U	SMPL
O4693-02 - 1	S-1(0-5)	86173	12.4455	-	7471B	PPB	05 Oct 2023 15:48:05	U	SMPL
O4693-03 - 1	S-2(10-15)	22999	3.3185	-	7471B	PPB	05 Oct 2023 15:50:21	U	SMPL
O4693-04 - 1	S-3(0-5)	80452	11.6190	-	7471B	PPB	05 Oct 2023 15:52:38	U	SMPL
O4704-01 - 1	S-4(0-5)	14209	2.0486	-	7471B	PPB	05 Oct 2023 15:54:56	U	SMPL
O4704-02 - 1	R-1(10-15)	19060	2.7495	-	7471B	PPB	05 Oct 2023 15:57:14	U	SMPL
O4704-03 - 1	R-2(5-10)	43695	6.3086	-	7471B	PPB	05 Oct 2023 15:59:30	U	SMPL
O4704-04 - 1	R-3(10-15)	47966	6.9256	-	7471B	PPB	05 Oct 2023 16:01:47	U	SMPL
O4704-05 - 1	SB-104-1(5-10)	21285	3.0709	-	7471B	PPB	05 Oct 2023 16:04:03	U	SMPL
CCV29 - 1	CCV29	36501	5.2692	-	7471B	PPB	05 Oct 2023 16:06:18	U	SMPL
CCB29 - 1	CCB29	57	0.0040	-	7471B	PPB	05 Oct 2023 16:08:34	U	SMPL
O4704-06 - 1	R-4(5-10)	17026	2.4556	-	7471B	PPB	05 Oct 2023 16:10:50	U	SMPL
O4720-01 - 1	ORA-1042	78	0.0071	-	7471B	PPB	05 Oct 2023 16:13:07	U	SMPL
O4731-01 - 1	DTP-1	1656	0.2350	-	7471B	PPB	05 Oct 2023 16:15:23	U	SMPL
O4731-04 - 1	DTP-2	773	0.1075	-	7471B	PPB	05 Oct 2023 16:17:39	U	SMPL
O4731-04DUP - 1	DTP-2DUP	579	0.0794	-	7471B	PPB	05 Oct 2023 16:19:54	U	SMPL
O4731-04MS - 1	DTP-2MS	30725	4.4347	-	7471B	PPB	05 Oct 2023 16:22:10	U	SMPL
O4731-04MSD - 1	DTP-2MSD	29926	4.3193	-	7471B	PPB	05 Oct 2023 16:24:27	U	SMPL
O4731-07 - 1	DTP-3	3656	0.5240	-	7471B	PPB	05 Oct 2023 16:26:44	U	SMPL
PB156160BL - 1	PBS	28	-0.0002	-	7471B	PPB	05 Oct 2023 16:29:01	U	SMPL
PB156160BS - 1	LCSS	28778	4.1535	-	7471B	PPB	05 Oct 2023 16:42:24	U	SMPL
CCV30 - 1	CCV30	37331	5.3891	-	7471B	PPB	05 Oct 2023 16:44:41	U	SMPL
CCB30 - 1	CCB30	41	0.0017	-	7471B	PPB	05 Oct 2023 16:46:56	U	SMPL
O4735-02 - 1	RT2662	7731	1.1127	-	7471B	PPB	05 Oct 2023 16:54:14	U	SMPL
O4735-02DUP - 1	RT2662DUP	8173	1.1766	-	7471B	PPB	05 Oct 2023 16:56:30	U	SMPL
O4735-02MS - 1	RT2662MS	31987	4.6171	-	7471B	PPB	05 Oct 2023 16:58:45	U	SMPL
O4735-02MSD - 1	RT2662MSD	32769	4.7300	-	7471B	PPB	05 Oct 2023 17:01:00	U	SMPL
O4736-02 - 1	72-11991	3704	0.5309	-	7471B	PPB	05 Oct 2023 17:03:16	U	SMPL
O4736-04 - 1	72-11939	2779	0.3973	-	7471B	PPB	05 Oct 2023 17:05:32	U	SMPL
O4753-01 - 1	TP-4	3034	0.4341	-	7471B	PPB	05 Oct 2023 17:07:48	U	SMPL
O4763-01 - 1	PAT-1-100423	6619	0.9521	-	7471B	PPB	05 Oct 2023 17:10:05	U	SMPL
O4765-01 - 1	TT-1	5853	0.8414	-	7471B	PPB	05 Oct 2023 17:12:21	U	SMPL
O4765-03 - 1	TT-2	2212	0.3154	-	7471B	PPB	05 Oct 2023 17:14:37	U	SMPL
CCV31 - 1	CCV31	34792	5.0223	-	7471B	PPB	05 Oct 2023 17:16:52	U	SMPL
CCB31 - 1	CCB31	77	0.0069	-	7471B	PPB	05 Oct 2023 17:19:10	U	SMPL
O4731-04LX5 - 1		179	0.0217	-	7471B	PPB	05 Oct 2023 17:21:26	U	SMPL
O4731-04A - 1		30005	4.3307	-	7471B	PPB	05 Oct 2023 17:23:41	U	SMPL
O4735-02LX5 - 1		1592	0.2258	-	7471B	PPB	05 Oct 2023 17:29:41	U	SMPL
O4735-02A - 1		34736	5.0142	-	7471B	PPB	05 Oct 2023 17:31:56	U	SMPL
O4693-02DLX2 - 1		43856	6.3318	-	7471B	PPB	05 Oct 2023 17:34:11	U	SMPL
O4693-04DLX2 - 1		41690	6.0189	-	7471B	PPB	05 Oct 2023 17:36:28	U	SMPL
CCV32 - 1	CCV32	35075	5.0632	-	7471B	PPB	05 Oct 2023 17:38:44	U	SMPL
CCB32 - 1	CCB32	89	0.0086	-	7471B	PPB	05 Oct 2023 17:41:02	U	SMPL