



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:45
In Date: 10/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 10/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4710-01	MJUBN5	1	1.15	8.57	9.72	7.7	76.4	
O4710-02	MJUBN6	2	1.19	8.41	9.6	7.24	71.9	
O4710-03	MJUBN7	3	1.15	8.81	9.96	7.27	69.5	
O4710-04	MJUBN8	4	1.16	8.80	9.96	8.26	80.7	
O4710-05	MJUBN9	5	1.16	8.52	9.68	7.22	71.1	
O4710-06	MJUBP0	6	1.15	8.77	9.92	7.82	76.1	
O4710-07	MJUBP1	7	1.19	8.67	9.86	8.84	88.2	
O4710-08	MJUBP3	8	1.15	8.30	9.45	8.87	93.0	
O4710-09	MJUBP4	9	1.13	8.55	9.68	8.9	90.9	
O4710-10	MJUBP5	10	1.14	8.53	9.67	7.91	79.4	
O4710-11	MJUBP7	11	1.16	8.72	9.88	6.81	64.8	
O4710-12	MJUBP9	12	1.14	8.45	9.59	7.44	74.6	
O4710-13	MJUBQ0	13	1.13	8.71	9.84	7.47	72.8	
O4710-14	MJUBQ1	14	1.16	8.80	9.96	7.61	73.3	
O4710-15	MJUBQ2	15	1.15	8.40	9.55	6.88	68.2	
O4710-16	MJUBQ2D	16	1.15	8.40	9.55	6.88	68.2	
O4710-17	MJUBQ2S	17	1.15	8.40	9.55	6.88	68.2	
O4710-18	MJUBQ4	18	1.19	8.54	9.73	7.29	71.4	
O4710-19	MJUBQ5	19	1.14	8.83	9.97	7.42	71.1	
O4710-20	MJUBQ6	20	1.16	8.54	9.7	8.39	84.7	
O4710-21	MJUBQ7	21	1.13	8.45	9.58	8.08	82.2	
O4710-22	MJUBQ8	22	1.13	8.48	9.61	8.4	85.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$