



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:25
In Date: 10/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127636

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4711-01	MJUBM4	1	1.19	8.61	9.8	9.11	92.0	
O4711-02	MJUBM5	2	1.13	8.80	9.93	8.8	87.2	
O4711-03	MJUBM6	3	1.15	8.76	9.91	7.19	68.9	
O4711-04	MJUBM7	4	1.15	8.84	9.99	7.91	76.5	
O4711-05	MJUBM8	5	1.13	8.64	9.77	8.94	90.4	
O4711-06	MJUBM9	6	1.19	8.48	9.67	8.18	82.4	
O4711-07	MJUBN0	7	1.15	8.56	9.71	7.87	78.5	
O4711-08	MJUBN1	8	1.12	8.76	9.88	7.7	75.1	
O4711-09	MJUBN2	9	1.15	8.39	9.54	7.38	74.3	
O4711-10	MJUBN3	10	1.13	8.73	9.86	9.19	92.3	
O4711-11	MJUBN4	11	1.17	8.60	9.77	8.75	88.1	
O4711-12	MJUBQ9	12	1.16	8.55	9.71	8.46	85.4	
O4711-13	MJUBR0	13	1.15	8.05	9.2	8.15	87.0	
O4711-14	MJUBR1	14	1.18	8.50	9.68	7.6	75.5	
O4711-15	MJUBR2	15	1.15	8.37	9.52	8.04	82.3	
O4711-16	MJUBR3	16	1.14	8.75	9.89	8.38	82.7	
O4711-17	MJUBR4	17	1.19	8.50	9.69	8.07	80.9	
O4711-18	MJUBR5	18	1.13	8.38	9.51	7.84	80.1	
O4711-19	MJUBR6	19	1.16	8.62	9.78	9.39	95.5	
O4711-20	MJUBR7	20	1.13	8.80	9.93	9.55	95.7	
O4711-21	MJUBR7D	21	1.13	8.80	9.93	9.55	95.7	
O4711-22	MJUBR7S	22	1.13	8.80	9.93	9.55	95.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$