

LB127642


7470A

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

52381



A= 0.0000e+000

B= 1.9143e-004

C= -1.6017e-001

Rho= 0.9997067

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.015 | 0.015  | 913   | 0.000      | 913   | 0     |       |       |       |     |
| 0.2    | 0.200  | 0.162 | -0.038 | 1682  | 0.0 %      | 1682  |       |       |       |       | -19 |
| 2.5    | 2.500  | 2.448 | -0.052 | 13626 | 0.0 %      | 13626 |       |       |       |       | -2  |
| 5.0    | 5.000  | 5.058 | 0.058  | 27256 | 0.0 %      | 27256 |       |       |       |       | 1   |
| 7.5    | 7.500  | 7.650 | 0.150  | 40801 | 0.0 %      | 40801 |       |       |       |       | 2   |
| 10.0   | 10.000 | 9.867 | -0.133 | 52381 | 0.0 %      | 52381 |       |       |       |       | -1  |

## LB127642 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Oct 2023 08:07:44

| Sample ID       | Extended ID | $\mu$ Abs. | Conc.   | Std Conc | Method | Units                | Date                 | Type | Type |
|-----------------|-------------|------------|---------|----------|--------|----------------------|----------------------|------|------|
| 0.0 - 1         | 50          | 913        | -       | 0.0000   | 7470A  | PPB                  | 06 Oct 2023 08:20:53 | S    | Std  |
| 0.2 - 1         | 50.2        | 1682       | -       | 0.2000   | 7470A  | PPB                  | 06 Oct 2023 08:23:10 | S    | Std  |
| 2.5 - 1         | 525         | 13626      | -       | 2.5000   | 7470A  | PPB                  | 06 Oct 2023 08:25:27 | S    | Std  |
| 5.0 - 1         | 55          | 27256      | -       | 5.0000   | 7470A  | PPB                  | 06 Oct 2023 08:27:43 | S    | Std  |
| 7.5 - 1         | 57.5        | 40801      | -       | 7.5000   | 7470A  | PPB                  | 06 Oct 2023 08:29:59 | S    | Std  |
| 10.0 - 1        | 50          | 52381      | -       | 10.0000  | 7470A  | PPB                  | 06 Oct 2023 08:32:15 | S    | Std  |
| ICV98 - 1       | ICV98       | 20883      | 3.8375  | -7470A   | PPB    | 06 Oct 2023 08:35:15 | U                    | SMPL |      |
| ICB98 - 1       | ICB98       | 432        | -0.0775 | -7470A   | PPB    | 06 Oct 2023 08:37:29 | U                    | SMPL |      |
| CCV32 - 1       | CCV32       | 27432      | 5.0912  | -7470A   | PPB    | 06 Oct 2023 08:39:44 | U                    | SMPL |      |
| CCB32 - 1       | CCB32       | 459        | -0.0723 | -7470A   | PPB    | 06 Oct 2023 08:41:59 | U                    | SMPL |      |
| CRA - 1         | CRA         | 1711       | 0.1674  | -7470A   | PPB    | 06 Oct 2023 08:49:24 | U                    | SMPL |      |
| HighStd - 1     | HighStd     | 52614      | 9.9119  | -7470A   | PPB    | 06 Oct 2023 08:51:39 | U                    | SMPL |      |
| ChkStd - 1      | ChkStd      | 38745      | 7.2569  | -7470A   | PPB    | 06 Oct 2023 08:53:54 | U                    | SMPL |      |
| PB156157BL - 1  | PBW         | 785        | -0.0099 | -7470A   | PPB    | 06 Oct 2023 08:58:43 | U                    | SMPL |      |
| PB156157BS - 1  | LCSW        | 19616      | 3.5950  | -7470A   | PPB    | 06 Oct 2023 09:00:59 | U                    | SMPL |      |
| O4712-03 - 1    | WC-1        | 135        | -0.1343 | -7470A   | PPB    | 06 Oct 2023 09:03:15 | U                    | SMPL |      |
| O4712-06 - 1    | WC-2        | 183        | -0.1251 | -7470A   | PPB    | 06 Oct 2023 09:05:31 | U                    | SMPL |      |
| O4714-03 - 1    | TP-1        | 156        | -0.1303 | -7470A   | PPB    | 06 Oct 2023 09:07:48 | U                    | SMPL |      |
| O4714-06 - 1    | TP-2        | 180        | -0.1257 | -7470A   | PPB    | 06 Oct 2023 09:10:04 | U                    | SMPL |      |
| O4714-09 - 1    | TP-3        | 191        | -0.1236 | -7470A   | PPB    | 06 Oct 2023 09:12:19 | U                    | SMPL |      |
| CCV33 - 1       | CCV33       | 27402      | 5.0855  | -7470A   | PPB    | 06 Oct 2023 09:14:34 | U                    | SMPL |      |
| CCB33 - 1       | CCB33       | 442        | -0.0756 | -7470A   | PPB    | 06 Oct 2023 09:16:49 | U                    | SMPL |      |
| O4714-12 - 1    | TP-4        | 187        | -0.1244 | -7470A   | PPB    | 06 Oct 2023 09:19:04 | U                    | SMPL |      |
| O4714-12DUP - 1 | TP-4DUP     | 176        | -0.1265 | -7470A   | PPB    | 06 Oct 2023 09:21:19 | U                    | SMPL |      |
| O4714-12MS - 1  | TP-4MS      | 21943      | 4.0404  | -7470A   | PPB    | 06 Oct 2023 09:23:34 | U                    | SMPL |      |
| O4714-12MSD - 1 | TP-4MSD     | 22740      | 4.1930  | -7470A   | PPB    | 06 Oct 2023 09:25:50 | U                    | SMPL |      |
| PB156164BL - 1  | PBW         | 794        | -0.0082 | -7470A   | PPB    | 06 Oct 2023 09:28:06 | U                    | SMPL |      |
| PB156164BS - 1  | LCSW        | 20082      | 3.6842  | -7470A   | PPB    | 06 Oct 2023 09:33:24 | U                    | SMPL |      |
| O4718-01 - 1    | A455        | 264        | -0.1096 | -7470A   | PPB    | 06 Oct 2023 09:35:40 | U                    | SMPL |      |
| O4718-01DUP - 1 | A455DUP     | 303        | -0.1022 | -7470A   | PPB    | 06 Oct 2023 09:37:56 | U                    | SMPL |      |
| O4718-01MS - 1  | A455MS      | 20724      | 3.8071  | -7470A   | PPB    | 06 Oct 2023 09:40:13 | U                    | SMPL |      |
| O4718-01MSD - 1 | A455MSD     | 20055      | 3.6790  | -7470A   | PPB    | 06 Oct 2023 09:42:28 | U                    | SMPL |      |
| CCV34 - 1       | CCV34       | 27225      | 5.0516  | -7470A   | PPB    | 06 Oct 2023 09:44:43 | U                    | SMPL |      |
| CCB34 - 1       | CCB34       | 428        | -0.0782 | -7470A   | PPB    | 06 Oct 2023 09:46:59 | U                    | SMPL |      |
| O4738-02 - 1    | COMP        | 1201       | 0.0697  | -7470A   | PPB    | 06 Oct 2023 09:49:14 | U                    | SMPL |      |
| O4767-04 - 1    | EB-1        | 690        | -0.0281 | -7470A   | PPB    | 06 Oct 2023 09:51:30 | U                    | SMPL |      |
| O4767-05 - 1    | EB-2        | 724        | -0.0216 | -7470A   | PPB    | 06 Oct 2023 09:53:45 | U                    | SMPL |      |
| PB156086TB - 1  | PB156086TB  | 728        | -0.0208 | -7470A   | PPB    | 06 Oct 2023 09:58:36 | U                    | SMPL |      |
| O4714-12LX5 - 1 |             | 65         | -0.1477 | -7470A   | PPB    | 06 Oct 2023 10:00:51 | U                    | SMPL |      |
| O4714-12A - 1   |             | 21601      | 3.9750  | -7470A   | PPB    | 06 Oct 2023 10:03:07 | U                    | SMPL |      |
| O4718-01LX5 - 1 |             | 41         | -0.1523 | -7470A   | PPB    | 06 Oct 2023 10:05:24 | U                    | SMPL |      |
| O4718-01A - 1   |             | 21827      | 4.0182  | -7470A   | PPB    | 06 Oct 2023 10:07:40 | U                    | SMPL |      |
| CCV35 - 1       | CCV35       | 27330      | 5.0717  | -7470A   | PPB    | 06 Oct 2023 10:09:57 | U                    | SMPL |      |
| CCB35 - 1       | CCB35       | 773        | -0.0122 | -7470A   | PPB    | 06 Oct 2023 10:12:13 | U                    | SMPL |      |