



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127643

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/06/2023	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	10/06/2023	08:16
3	CAL3	1	Water	NA	NA	20.5	10.02	10/06/2023	08:19
4	ICV	1	Water	NA	NA	20.6	7.01	10/06/2023	08:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/06/2023	08:23
6	O4718-01	1	Water	NA	NA	22.2	11.13	10/06/2023	08:30
7	O4770-02	1	Water	NA	NA	20.6	6.93	10/06/2023	08:35
8	O4770-04	1	Water	NA	NA	21.6	6.89	10/06/2023	08:40
9	O4770-06	1	Water	NA	NA	21.6	7.67	10/06/2023	08:45
10	O4778-10	1	Water	NA	NA	22.6	7.00	10/06/2023	08:55
11	O4778-10DUP	1	Water	NA	NA	22.7	7.01	10/06/2023	08:56
12	CCV2	1	Water	NA	NA	20.6	12.02	10/06/2023	09:00