



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/9/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:25
In Date: 10/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 10/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127646

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4755-01	YC5M2	1	1.14	8.78	9.92	9.43	94.4	
O4755-02	YC5M2D	2	1.14	8.78	9.92	9.43	94.4	
O4755-03	YC5M2S	3	1.14	8.78	9.92	9.43	94.4	
O4755-04	YC5M3	4	1.14	8.36	9.5	8.84	92.1	
O4755-05	YC5M4	5	1.16	8.76	9.92	9.7	97.5	
O4755-06	YC5M5	6	1.19	8.52	9.71	9.33	95.5	
O4755-07	YC5M6	7	1.17	8.60	9.77	9.34	95.0	
O4755-08	YC5M7	8	1.19	8.43	9.62	9.17	94.7	
O4755-09	YC5M8	9	1.15	8.39	9.54	9.35	97.7	
O4755-10	YC5M9	10	1.19	8.42	9.61	9.21	95.2	
O4755-11	YC5N0	11	1.19	8.60	9.79	8.96	90.3	
O4755-12	YC5N1	12	1.16	8.31	9.47	8.66	90.3	
O4755-13	YC5N2	13	1.15	8.40	9.55	9.11	94.8	
O4755-14	YC5N3	14	1.19	8.40	9.59	8.94	92.3	
O4755-15	YC5N4	15	1.15	8.43	9.58	9.21	95.6	
O4755-16	YC5N5	16	1.17	8.39	9.56	9.14	95.0	
O4755-17	YC5N6	17	1.15	8.43	9.58	8.98	92.9	
O4755-18	YC5N7	18	1.13	8.38	9.51	9.06	94.6	
O4755-19	YC5N8	19	1.19	8.42	9.61	8.89	91.4	
O4755-20	YC5N9	20	1.19	8.54	9.73	8.82	89.3	
O4755-21	YC5P2	21	1.15	8.49	9.64	9.26	95.5	
O4755-22	YC5P9	22	1.19	8.44	9.63	9.38	97.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$