



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/9/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 10/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127647

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4756-01	YC5P0	1	1.19	8.44	9.63	9.44	97.7	
O4756-02	YC5P1	2	1.19	8.43	9.62	9.24	95.5	
O4756-03	YC5P3	3	1.17	8.60	9.77	9.3	94.5	
O4756-04	YC5P4	4	1.13	8.47	9.6	9.41	97.8	
O4756-05	YC5P5	5	1.16	8.50	9.66	9.18	94.4	
O4756-06	YC5P6	6	1.14	8.69	9.83	9.64	97.8	
O4756-07	YC5P7	7	1.16	8.50	9.66	9.33	96.1	
O4756-08	YC5P8	8	1.13	8.71	9.84	9.36	94.5	
O4756-09	YC5Q0	9	1.16	8.68	9.84	9.32	94.0	
O4756-10	YC5Y8	10	1.19	8.54	9.73	9.28	94.7	
O4756-11	YC5Q1	11	1.15	8.37	9.52	8.96	93.3	
O4756-12	YC5Q2	12	1.19	8.42	9.61	8.98	92.5	
O4756-13	YC5Q3	13	1.13	8.45	9.58	9.33	97.0	
O4756-14	YC5Q4	14	1.19	8.52	9.71	9.18	93.8	
O4756-15	YC5Q5	15	1.15	8.44	9.59	9.34	97.0	
O4756-16	YC5Q6	16	1.19	8.52	9.71	9.2	94.0	
O4756-17	YC5Q6D	17	1.19	8.52	9.71	9.2	94.0	
O4756-18	YC5Q6S	18	1.19	8.52	9.71	9.2	94.0	
O4756-19	YC5Q7	19	1.18	8.55	9.73	9.36	95.7	
O4756-20	YC5R4	20	1.19	8.42	9.61	8.88	91.3	
O4756-21	YC5R7	21	1.17	8.60	9.77	9.11	92.3	
O4756-22	YC5R8	22	1.15	8.67	9.82	9.17	92.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$