


LB129648

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

61991



Linear

A= 0.0000e+000

B= 1.5903e-004 slope

C= -1.8104e-002 y-intercept

Rho= 0.9996942

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/10
0.00	0.000	-0.008	-0.008	63	0.000	63					
0.05	0.050	0.041	-0.009	372	0.0 %	372					-18
0.20	0.200	0.147	-0.053	1039	0.0 %	1039	0				-27
2.50	2.500	2.492	-0.008	15782	0.0 %	15782					0
5.00	5.000	5.086	0.086	32094	0.0 %	32094					2
7.50	7.500	7.652	0.152	48231	0.0 %	48231					2
10.0	10.000	9.840	-0.160	61991	0.0 %	61991					-2

LB127648
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	63	-		0 SFAM01.1	PPB	10/6/2023 10:47	Std
	0.05 S01	372	-		0.05 SFAM01.1	PPB	10/6/2023 10:49	Std
	0.2 S02	1039	-		0.2 SFAM01.1	PPB	10/6/2023 10:52	Std
	2.5 S03	15782	-		2.5 SFAM01.1	PPB	10/6/2023 10:54	Std
	5 S04	32094	-		5 SFAM01.1	PPB	10/6/2023 10:56	Std
	7.5 S05	48231	-		7.5 SFAM01.1	PPB	10/6/2023 10:58	Std
	10 S06	61991	-		10 SFAM01.1	PPB	10/6/2023 11:03	Std
ICV099	ICV099	25725	4.0729 -		SFAM01.1	PPB	10/6/2023 11:06	SMPL
ICB099	ICB099	14	-0.0159 -		SFAM01.1	PPB	10/6/2023 11:09	SMPL
CCV036	CCV036	31642	5.0139 -		SFAM01.1	PPB	10/6/2023 11:11	SMPL
CCB036	CCB036	32	-0.013 -		SFAM01.1	PPB	10/6/2023 11:13	SMPL
PB156165BL	PBS165	104	-0.0016 -		SFAM01.1	PPB	10/6/2023 11:15	SMPL
O4532-01	MJKW40	818	0.112 -		SFAM01.1	PPB	10/6/2023 11:18	SMPL
O4532-02	MJKW41	1372	0.2001 -		SFAM01.1	PPB	10/6/2023 11:20	SMPL
O4532-03	MJKW42	816	0.1117 -		SFAM01.1	PPB	10/6/2023 11:22	SMPL
O4532-04	MJKW43	577	0.0737 -		SFAM01.1	PPB	10/6/2023 11:24	SMPL
O4532-05	MJKW44	886	0.1228 -		SFAM01.1	PPB	10/6/2023 11:27	SMPL
O4532-06	MJKW46	736	0.0989 -		SFAM01.1	PPB	10/6/2023 11:29	SMPL
O4532-07	MJKW47	6030	0.9408 -		SFAM01.1	PPB	10/6/2023 11:31	SMPL
O4532-08	MJKW48	678	0.0897 -		SFAM01.1	PPB	10/6/2023 11:33	SMPL
O4532-09	MJKW49	511	0.0632 -		SFAM01.1	PPB	10/6/2023 11:36	SMPL
O4532-10	MJKW82	1695	0.2515 -		SFAM01.1	PPB	10/6/2023 11:38	SMPL
O4532-11	MJKW83	896	0.1244 -		SFAM01.1	PPB	10/6/2023 11:40	SMPL
O4532-12	MJKW84	628	0.0818 -		SFAM01.1	PPB	10/6/2023 11:42	SMPL
O4532-13	MJKW85	678	0.0897 -		SFAM01.1	PPB	10/6/2023 11:45	SMPL
O4532-14	MJKW86	2126	0.32 -		SFAM01.1	PPB	10/6/2023 11:47	SMPL
O4532-15	MJKW87	489	0.0597 -		SFAM01.1	PPB	10/6/2023 11:49	SMPL
O4532-16	MJKW88	490	0.0598 -		SFAM01.1	PPB	10/6/2023 11:52	SMPL
O4532-17	MJKW89	681	0.0902 -		SFAM01.1	PPB	10/6/2023 11:54	SMPL
O4532-18	MJKW90	1334	0.194 -		SFAM01.1	PPB	10/6/2023 11:56	SMPL
O4532-19	MJKW91	2496	0.3788 -		SFAM01.1	PPB	10/6/2023 11:58	SMPL
CCV037	CCV037	31915	5.0573 -		SFAM01.1	PPB	10/6/2023 12:01	SMPL
CCB037	CCB037	28	-0.0137 -		SFAM01.1	PPB	10/6/2023 12:03	SMPL
O4532-20	MJKW92	966	0.1355 -		SFAM01.1	PPB	10/6/2023 12:05	SMPL
O4532-21	MJKW92D	815	0.1115 -		SFAM01.1	PPB	10/6/2023 12:07	SMPL
O4532-22	MJKW92S	15780	2.4914 -		SFAM01.1	PPB	10/6/2023 12:10	SMPL
PB156166BL	PBS166	114	0 -		SFAM01.1	PPB	10/6/2023 12:12	SMPL
O4533-01	MJKW60	1301	0.1888 -		SFAM01.1	PPB	10/6/2023 12:14	SMPL
O4533-02	MJKW61	708	0.0945 -		SFAM01.1	PPB	10/6/2023 12:16	SMPL
O4533-03	MJKW62	867	0.1198 -		SFAM01.1	PPB	10/6/2023 12:19	SMPL
O4533-04	MJKW63	627	0.0816 -		SFAM01.1	PPB	10/6/2023 12:21	SMPL
O4533-05	MJKW64	557	0.0705 -		SFAM01.1	PPB	10/6/2023 12:23	SMPL
O4533-06	MJKW65	664	0.0875 -		SFAM01.1	PPB	10/6/2023 12:25	SMPL
O4533-07	MJKW66	662	0.0872 -		SFAM01.1	PPB	10/6/2023 12:28	SMPL
O4533-08	MJKW66D	797	0.1086 -		SFAM01.1	PPB	10/6/2023 12:30	SMPL
O4533-09	MJKW66S	15796	2.4939 -		SFAM01.1	PPB	10/6/2023 12:32	SMPL

LB127648
INSTRUMENT ID : CV1

O4533-10	MJKW67	664	0.0875 -	SFAM01.1	PPB	10/6/2023 12:35 SMPL
O4533-11	MJKW68	754	0.1018 -	SFAM01.1	PPB	10/6/2023 12:37 SMPL
O4533-12	MJKW69	767	0.1039 -	SFAM01.1	PPB	10/6/2023 12:39 SMPL
O4533-13	MJUBG6	359	0.039 -	SFAM01.1	PPB	10/6/2023 12:41 SMPL
O4533-14	MJUBG7	1807	0.2693 -	SFAM01.1	PPB	10/6/2023 12:44 SMPL
O4533-15	MJUBG8	824	0.1129 -	SFAM01.1	PPB	10/6/2023 12:46 SMPL
O4533-16	MJUBG9	768	0.104 -	SFAM01.1	PPB	10/6/2023 12:48 SMPL
CCV038	CCV038	31113	4.9298 -	SFAM01.1	PPB	10/6/2023 12:50 SMPL
CCB038	CCB038	18	-0.0152 -	SFAM01.1	PPB	10/6/2023 12:53 SMPL
O4533-17	MJUBH0	2754	0.4199 -	SFAM01.1	PPB	10/6/2023 12:55 SMPL
O4533-18	MJUBH1	1149	0.1646 -	SFAM01.1	PPB	10/6/2023 12:57 SMPL
O4533-19	MJUBH2	820	0.1123 -	SFAM01.1	PPB	10/6/2023 12:59 SMPL
O4533-20	MJUBH3	808	0.1104 -	SFAM01.1	PPB	10/6/2023 13:02 SMPL
O4533-21	MJUB47	3658	0.5636 -	SFAM01.1	PPB	10/6/2023 13:04 SMPL
O4533-22	MJUB48	6123	0.9556 -	SFAM01.1	PPB	10/6/2023 13:06 SMPL
PB156167BL	PBS167	180	0.0105 -	SFAM01.1	PPB	10/6/2023 13:10 SMPL
O4534-01	MJKW50	1864	0.2783 -	SFAM01.1	PPB	10/6/2023 13:13 SMPL
O4534-02	MJKW51	1773	0.2639 -	SFAM01.1	PPB	10/6/2023 13:15 SMPL
O4534-03	MJKW52	1698	0.2519 -	SFAM01.1	PPB	10/6/2023 13:17 SMPL
O4534-04	MJKW53	2539	0.3857 -	SFAM01.1	PPB	10/6/2023 13:20 SMPL
O4534-05	MJKW54	1520	0.2236 -	SFAM01.1	PPB	10/6/2023 13:22 SMPL
O4534-06	MJKW55	1775	0.2642 -	SFAM01.1	PPB	10/6/2023 13:24 SMPL
O4534-07	MJKW56	1296	0.188 -	SFAM01.1	PPB	10/6/2023 13:26 SMPL
O4534-08	MJKW57	1114	0.1591 -	SFAM01.1	PPB	10/6/2023 13:29 SMPL
O4534-09	MJKW58	854	0.1177 -	SFAM01.1	PPB	10/6/2023 13:31 SMPL
O4534-10	MJKW59	1115	0.1592 -	SFAM01.1	PPB	10/6/2023 13:33 SMPL
O4534-11	MJKWA7	4171	0.6452 -	SFAM01.1	PPB	10/6/2023 13:35 SMPL
O4534-12	MJKWA8	2732	0.4164 -	SFAM01.1	PPB	10/6/2023 13:38 SMPL
O4534-13	MJKWA9	1106	0.1578 -	SFAM01.1	PPB	10/6/2023 13:40 SMPL
CCV039	CCV039	31195	4.9428 -	SFAM01.1	PPB	10/6/2023 13:42 SMPL
CCB039	CCB039	27	-0.0138 -	SFAM01.1	PPB	10/6/2023 13:44 SMPL
O4534-14	MJKWB0	3153	0.4833 -	SFAM01.1	PPB	10/6/2023 13:47 SMPL
O4534-15	MJKWB1	603	0.0778 -	SFAM01.1	PPB	10/6/2023 13:49 SMPL
O4534-16	MJKWB2	556	0.0703 -	SFAM01.1	PPB	10/6/2023 13:51 SMPL
O4534-17	MJKWB3	455	0.0543 -	SFAM01.1	PPB	10/6/2023 13:54 SMPL
O4534-18	MJKWB3D	508	0.0627 -	SFAM01.1	PPB	10/6/2023 13:56 SMPL
O4534-19	MJKWB3S	16751	2.6458 -	SFAM01.1	PPB	10/6/2023 13:58 SMPL
PB156168BL	PBS168	153	0.0062 -	SFAM01.1	PPB	10/6/2023 14:00 SMPL
O4552-01	MJKW45	619	0.0803 -	SFAM01.1	PPB	10/6/2023 14:03 SMPL
O4552-02	MJKW45D	680	0.09 -	SFAM01.1	PPB	10/6/2023 14:05 SMPL
O4552-03	MJKW45S	16236	2.5639 -	SFAM01.1	PPB	10/6/2023 14:07 SMPL
PB156169BL	PBS169	162	0.0077 -	SFAM01.1	PPB	10/6/2023 14:09 SMPL
O4553-01	MJUB49	7477	1.171 -	SFAM01.1	PPB	10/6/2023 14:12 SMPL
O4553-02	MJUB50	1084	0.1543 -	SFAM01.1	PPB	10/6/2023 14:14 SMPL
O4553-03	MJUB51	1007	0.142 -	SFAM01.1	PPB	10/6/2023 14:16 SMPL
O4553-04	MJUB52	933	0.1303 -	SFAM01.1	PPB	10/6/2023 14:19 SMPL
O4553-05	MJUB54	824	0.1129 -	SFAM01.1	PPB	10/6/2023 14:21 SMPL

LB127648

INSTRUMENT ID : CV1

O4553-06	MJUB55	1019	0.1439 -	SFAM01.1	PPB	10/6/2023 14:23 SMPL
O4553-07	MJUB56	1305	0.1894 -	SFAM01.1	PPB	10/6/2023 14:25 SMPL
O4553-08	MJUB56D	1112	0.1587 -	SFAM01.1	PPB	10/6/2023 14:28 SMPL
O4553-09	MJUB56S	16796	2.653 -	SFAM01.1	PPB	10/6/2023 14:30 SMPL
CCV040	CCV040	31930	5.0597 -	SFAM01.1	PPB	10/6/2023 14:32 SMPL
CCB040	CCB040	26	-0.014 -	SFAM01.1	PPB	10/6/2023 14:35 SMPL