



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/9/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 10/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127649

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4757-01	YC5R9	1	1.15	8.43	9.58	9.03	93.5	
O4757-02	YC5S0	2	1.13	8.68	9.81	9.00	90.7	
O4757-03	YC5S0D	3	1.13	8.68	9.81	9.00	90.7	
O4757-04	YC5S0S	4	1.13	8.68	9.81	9.00	90.7	
O4757-05	YC5S1	5	1.19	8.63	9.82	9.4	95.1	
O4757-06	YC5S2	6	1.19	8.53	9.72	9.03	91.9	
O4757-07	YC5S3	7	1.16	8.55	9.71	9.47	97.2	
O4757-08	YC5S4	8	1.19	8.64	9.83	9.18	92.5	
O4757-09	YC5S5	9	1.15	8.42	9.57	9.36	97.5	
O4757-10	YC5S6	10	1.19	8.72	9.91	9.24	92.3	
O4757-11	YC5S7	11	1.19	8.64	9.83	9.34	94.3	
O4757-12	YC5Y7	12	1.19	8.62	9.81	9.55	97.0	
O4757-13	YC5Q8	13	1.15	8.38	9.53	9.16	95.6	
O4757-14	YC5Q9	14	1.15	8.36	9.51	9.2	96.3	
O4757-15	YC5R0	15	1.16	8.50	9.66	9.4	96.9	
O4757-16	YC5R1	16	1.14	8.34	9.48	9.18	96.4	
O4757-17	YC5R2	17	1.19	8.42	9.61	9.38	97.3	
O4757-18	YC5R3	18	1.16	8.47	9.63	9.36	96.8	
O4757-19	YC5R5	19	1.17	8.69	9.86	9.14	91.7	
O4757-20	YC5R6	20	1.13	8.66	9.79	9.48	96.4	
O4757-21	YC5S8	21	1.18	8.44	9.62	9.26	95.7	
O4757-22	YC5S9	22	1.18	8.57	9.75	9.2	93.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$