



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 10/06/2023 16:25 TEMP1 OUT: 103 °C 10/07/2023 07:30
TEMP2 IN: 104 °C 10/07/2023 08:00 TEMP2 OUT: 104 °C 10/07/2023 09:30
TEMP3 IN: 540 °C 10/07/2023 10:00 TEMP3 OUT: 540 °C 10/07/2023 11:30
TEMP4 IN: 560 °C 10/07/2023 12:00 TEMP4 OUT: 560 °C 10/07/2023 13:30

Run Number: LB127654

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB127654BL	89.6527	89.6527	100	89.6527	89.6527	89.6527	89.6527	0.0000	0
2	O4709-01	89.8774	89.8774	100	90.3272	90.3272	89.9928	89.9928	0.3344	3344
3	O4778-10	85.4469	85.4469	100	85.4971	85.4971	85.4746	85.4746	0.0225	225
4	O4778-10DUP	94.8770	94.8770	100	94.9281	94.9281	94.9061	94.9061	0.0220	220

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

W-127654

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS 04788 WATER

WorkList ID : 174352

Department : Wet-Chemistry

Date : 10-06-2023 13:02:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4709-01	55086-040423	Water	TVS	Cool 4 deg C	CHEM02	QA Of	10/03/2023	SM2540 G
O4778-10	☒ NBS-GROUNDWATER	Water	TVS	Cool 4 deg C	PSEG03	H11	10/05/2023	SM2540 G

Date/Time 13:30
Raw Sample Received by: JPL
Raw Sample Relinquished by: JPL

Date/Time 10/06/23 16:30
Raw Sample Received by: JPL
Raw Sample Relinquished by: JPL