

LB127656

SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

64464

A= 0.0000e+000

B= 1.5570e-004 slope

C= 2.0392e-002 y-intercept

Rho= 0.9999358

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/L
0.00	0.000	0.051	0.051	199	0.000	199					0.00
0.05	0.050					372					-10
0.20	0.200	0.181	-0.019	1031	0.0 %	1031	0				0
2.50	2.500	2.502	0.002	15937	0.0 %	15937					11
5.00	5.000	4.960	-0.040	31723	0.0 %	31723					11
7.50	7.500	7.449	-0.051	47714	0.0 %	47714					11
10.0	10.000	10.057	0.057	64464	0.0 %	64464					1

LB127656

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	199	-		0 SFAM01.1	PPB	10/6/2023 14:51	Std
	0.2 S01	1031	-		0.2 SFAM01.1	PPB	10/6/2023 14:55	Std
	2.5 S02	15937	-		2.5 SFAM01.1	PPB	10/6/2023 14:58	Std
	5 S03	31723	-		5 SFAM01.1	PPB	10/6/2023 15:00	Std
	7.5 S04	47714	-		7.5 SFAM01.1	PPB	10/6/2023 15:02	Std
	10 S05	64464	-		10 SFAM01.1	PPB	10/6/2023 15:08	Std
ICV001	ICV001	25502	3.991 -		SFAM01.1	PPB	10/6/2023 15:11	SMPL
ICB001	ICB001	253	0.0598 -		SFAM01.1	PPB	10/6/2023 15:13	SMPL
CCV041	CCV041	31855	4.9801 -		SFAM01.1	PPB	10/6/2023 15:16	SMPL
CCB041	CCB041	162	0.0456 -		SFAM01.1	PPB	10/6/2023 15:18	SMPL
PB156199BL	PBS199	370	0.078 -		SFAM01.1	PPB	10/6/2023 15:20	SMPL
O4561-01	MEZYL7	478	0.0948 -		SFAM01.1	PPB	10/6/2023 15:22	SMPL
O4561-02	MEZYL8	1172	0.2029 -		SFAM01.1	PPB	10/6/2023 15:25	SMPL
O4561-03	MEZYL9	1495	0.2532 -		SFAM01.1	PPB	10/6/2023 15:27	SMPL
O4561-04	MEZYM0	4087	0.6567 -		SFAM01.1	PPB	10/6/2023 15:29	SMPL
O4561-05	MEZYM1	1935	0.3217 -		SFAM01.1	PPB	10/6/2023 15:38	SMPL
O4561-06	MEZYM1D	1756	0.2938 -		SFAM01.1	PPB	10/6/2023 15:40	SMPL
O4561-07	MEZYM1S	18802	2.9478 -		SFAM01.1	PPB	10/6/2023 15:45	SMPL
O4561-08	MEZYM2	434	0.088 -		SFAM01.1	PPB	10/6/2023 15:47	SMPL
CCV042	CCV042	31966	4.9974 -		SFAM01.1	PPB	10/6/2023 15:49	SMPL
CCB042	CCB042	165	0.0461 -		SFAM01.1	PPB	10/6/2023 15:52	SMPL