



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127659

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	10/06/2023	13:10
2	O4770-02		1	12	100.00	>212.0	>212.0	10/06/2023	13:40
3	O4770-02DUP		1	12	100.00	>212.0	>212.0	10/06/2023	14:10
4	O4770-04		1	14	100.00	>212.0	>212.0	10/06/2023	14:40
5	O4770-06		1	12	100.00	>212.0	>212.0	10/06/2023	15:10
6	O4778-10		1	11	100.00	>212.0	>212.0	10/06/2023	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127659

WorkList Name : fp-10-06

WorkList ID : 174315

Department : Wet-Chemistry

Date : 10-06-2023 08:03:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4770-02 C	1065	Water	Flash Point	Cool 4 deg C	PSEG03	E22	10/05/2023	1010B
O4770-04 C	1057	Water	Flash Point	Cool 4 deg C	PSEG03	E22	10/05/2023	1010B
O4770-06 C	1058	Water	Flash Point	Cool 4 deg C	PSEG03	E22	10/05/2023	1010B
O4778-10 Q	NBS-GROUNDWATER	Water	Flash Point	Cool 4 deg C	PSEG03	H11	10/05/2023	1010B

Date/Time 10/06/2023 13.00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 10/06/2023 16.00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC