



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127662

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/06/2023	16:00
2	CAL2	1	Water	NA	NA	20.6	7.00	10/06/2023	16:02
3	CAL3	1	Water	NA	NA	20.6	10.02	10/06/2023	16:05
4	ICV	1	Water	NA	NA	20.5	7.00	10/06/2023	16:10
5	CCV1	1	Water	NA	NA	20.6	2.01	10/06/2023	16:11
6	O4798-01	1	Solid	20.02	20	23.0	9.29	10/06/2023	16:20
7	CCV2	1	Water	NA	NA	20.7	12.03	10/06/2023	16:22