

LA127669


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

58717



A= 0.0000e+000

B= 1.7054e-004 slope

C= -1.1608e-002 y-intercept

Rho= 0.9999987

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.008	0.008	115	0.000	115				
0.05	0.050					372	0			
0.20	0.200	0.191	-0.009	1187	0.0 %	1187				
2.50	2.500	2.505	0.005	14758	0.0 %	14758				
5.00	5.000	4.994	-0.006	29354	0.0 %	29354				
7.50	7.500	7.500	-0.000	44045	0.0 %	44045				
10.0	10.000	10.002	0.002	58717	0.0 %	58717				

0/000
44000

LB127663

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	115	-		0 SFAM01.1	PPB	10/6/2023 16:07	Std
	0.2	1187	-		0.2 SFAM01.1	PPB	10/6/2023 16:09	Std
	2.5	14758	-		2.5 SFAM01.1	PPB	10/6/2023 16:12	Std
	5	29354	-		5 SFAM01.1	PPB	10/6/2023 16:14	Std
	7.5	44045	-		7.5 SFAM01.1	PPB	10/6/2023 16:16	Std
	10	58717	-		10 SFAM01.1	PPB	10/6/2023 16:18	Std
ICV002	ICV002	24499	4.1664 -		SFAM01.1	PPB	10/6/2023 16:21	SMPL
ICB002	ICB002	49	-0.0033 -		SFAM01.1	PPB	10/6/2023 16:24	SMPL
CCV043	CCV043	29436	5.0084 -		SFAM01.1	PPB	10/6/2023 16:26	SMPL
CCB043	CCB043	40	-0.0048 -		SFAM01.1	PPB	10/6/2023 16:28	SMPL
PB156202BL	PBW202	96	0.0048 -		SFAM01.1	PPB	10/6/2023 16:30	SMPL
O4561-09	MEZYP6	598	0.0904 -		SFAM01.1	PPB	10/6/2023 16:33	SMPL
O4584-21	MJKWE0	1038	0.1654 -		SFAM01.1	PPB	10/6/2023 16:35	SMPL
CCV044	CCV044	29301	4.9853 -		SFAM01.1	PPB	10/6/2023 16:37	SMPL
CCB044	CCB044	39	-0.005 -		SFAM01.1	PPB	10/6/2023 16:39	SMPL