



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/9/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:25
In Date: 10/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 10/08/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127671

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4777-01	MJUBR8	1	1.13	8.60	9.73	9.29	94.9	
O4777-02	MJUBR9	2	1.15	8.73	9.88	8.63	85.7	
O4777-03	MJUBS0	3	1.17	8.44	9.61	8.5	86.8	
O4777-04	MJUBS1	4	1.15	8.48	9.63	8.79	90.1	
O4777-05	MJUBS2	5	1.13	8.46	9.59	7.17	71.4	
O4777-06	MJUBS3	6	1.16	8.50	9.66	7.18	70.8	
O4777-07	MJUBS4	7	1.18	8.34	9.52	7.24	72.7	
O4777-08	MJUBS5	8	1.14	8.37	9.51	7.14	71.7	
O4777-09	MJUBS6	9	1.15	8.50	9.65	7.35	72.9	
O4777-10	MJUBS7	10	1.19	8.39	9.58	7.74	78.1	
O4777-11	MJUBS8	11	1.19	8.63	9.82	7.67	75.1	
O4777-12	MJUBS9	12	1.15	8.51	9.66	7.76	77.7	
O4777-13	MJUBT0	13	1.15	8.41	9.56	8.07	82.3	
O4777-14	MJUBT1	14	1.15	8.36	9.51	8.07	82.8	
O4777-15	MJUBT2	15	1.19	8.71	9.9	9.47	95.1	
O4777-16	MJUBT3	16	1.19	8.69	9.88	9.51	95.7	
O4777-17	MJUBT4	17	1.16	8.81	9.97	9.66	96.5	
O4777-18	MJUBT5	18	1.16	8.45	9.61	8.64	88.5	
O4777-19	MJUBT5D	19	1.16	8.45	9.61	8.64	88.5	
O4777-20	MJUBT5S	20	1.16	8.45	9.61	8.64	88.5	
O4777-21	MJUBT6	21	1.19	8.47	9.66	9.12	93.6	
O4777-22	MJUBT7	22	1.16	8.82	9.98	9.55	95.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$