

LB129681


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54221



A= 0.0000e+000

B= 1.8443e-004

C= 4.5280e-002

Rho= 0.9998389

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	0.061	0.061	86	0.000	86	0				18
0.2	0.200	0.236	0.036	1036	0.0 %	1036					-2
2.5	2.500	2.444	-0.056	13006	0.0 %	13006					-2
5.0	5.000	4.878	-0.122	26205	0.0 %	26205					0
7.5	7.500	7.535	0.035	40612	0.0 %	40612					0
10.0	10.000	10.045	0.045	54221	0.0 %	54221					0

LB127681 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Oct 2023 11:32:12

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	86	-	0.0000	7470A	PPB	09 Oct 2023 11:38:46	S	Std
0.2 - 1	50.2	1036	-	0.2000	7470A	PPB	09 Oct 2023 11:41:03	S	Std
2.5 - 1	52.5	13008	-	2.5000	7470A	PPB	09 Oct 2023 11:43:20	S	Std
5.0 - 1	55	26205	-	5.0000	7470A	PPB	09 Oct 2023 11:45:36	S	Std
7.5 - 1	57.5	40612	-	7.5000	7470A	PPB	09 Oct 2023 11:47:53	S	Std
10.0 - 1	58.5	54221	-	10.0000	7470A	PPB	09 Oct 2023 11:50:08	S	Std
ICV03 - 1	ICV03	21269	3.9679	-7470A	PPB	09 Oct 2023 11:55:02	U	SMPL	
ICB03 - 1	ICB03	-35	0.0388	-7470A	PPB	09 Oct 2023 11:57:17	U	SMPL	
CCV45 - 1	CCV45	26226	4.8821	-7470A	PPB	09 Oct 2023 11:59:31	U	SMPL	
CCB45 - 1	CCB45	-18	0.0420	-7470A	PPB	09 Oct 2023 12:01:46	U	SMPL	
CRA - 1	CRA	979	0.2258	-7470A	PPB	09 Oct 2023 12:04:02	U	SMPL	
HighStd - 1	HighStd	53034	9.8262	-7470A	PPB	09 Oct 2023 12:06:17	U	SMPL	
ChkStd - 1	ChkStd	36945	6.8590	-7470A	PPB	09 Oct 2023 12:08:32	U	SMPL	
PB156203BL - 1	PBW	-59	0.0344	-7470A	PPB	09 Oct 2023 12:10:48	U	SMPL	
PB156203BS - 1	LCSW	22408	4.1779	-7470A	PPB	09 Oct 2023 12:13:05	U	SMPL	
O4731-03 - 1	DTP-1	-29	0.0399	-7470A	PPB	09 Oct 2023 12:15:21	U	SMPL	
O4731-06 - 1	DTP-2	29	0.0506	-7470A	PPB	09 Oct 2023 12:17:37	U	SMPL	
O4731-09 - 1	DTP-3	35	0.0517	-7470A	PPB	09 Oct 2023 12:19:54	U	SMPL	
O4733-02 - 1	20231004-WC1	61	0.0565	-7470A	PPB	09 Oct 2023 12:22:11	U	SMPL	
O4762-05 - 1	ROOSEVELT-BRIDGE-SANDBLAST-WEST	69	0.0580	-7470A	PPB	09 Oct 2023 12:24:26	U	SMPL	
CCV46 - 1	CCV46	25345	4.7196	-7470A	PPB	09 Oct 2023 12:26:41	U	SMPL	
CCB46 - 1	CCB46	-23	0.0410	-7470A	PPB	09 Oct 2023 12:28:56	U	SMPL	
O4762-06 - 1	ROOSEVELT-BRIDGE-SANDBLAST-EAST	37	0.0521	-7470A	PPB	09 Oct 2023 12:31:11	U	SMPL	
O4765-02 - 1	TT-1	41	0.0528	-7470A	PPB	09 Oct 2023 12:33:26	U	SMPL	
O4765-04 - 1	TT-2	39	0.0525	-7470A	PPB	09 Oct 2023 12:35:42	U	SMPL	
O4778-03 - 1	SOIL-1	43	0.0532	-7470A	PPB	09 Oct 2023 12:37:57	U	SMPL	
O4778-06 - 1	SOIL-2	44	0.0534	-7470A	PPB	09 Oct 2023 12:40:14	U	SMPL	
O4778-09 - 1	SOIL-3	66	0.0575	-7470A	PPB	09 Oct 2023 12:42:30	U	SMPL	
O4778-09DUP - 1	SOIL-3DUP	52	0.0549	-7470A	PPB	09 Oct 2023 12:44:46	U	SMPL	
O4778-09MS - 1	SOIL-3MS	22619	4.2168	-7470A	PPB	09 Oct 2023 12:47:03	U	SMPL	
O4778-09MSD - 1	SOIL-3MSD	23165	4.3175	-7470A	PPB	09 Oct 2023 12:49:20	U	SMPL	
CCV47 - 1	CCV47	24721	4.6045	-7470A	PPB	09 Oct 2023 12:51:36	U	SMPL	
CCB47 - 1	CCB47	-100	0.0268	-7470A	PPB	09 Oct 2023 12:53:52	U	SMPL	
PB156142TB - 1	PB156142TB	14	0.0479	-7470A	PPB	09 Oct 2023 12:56:09	U	SMPL	
O4778-09LX5 - 1		30	0.0508	-7470A	PPB	09 Oct 2023 12:58:24	U	SMPL	
O4778-09A - 1		20189	3.7687	-7470A	PPB	09 Oct 2023 13:00:39	U	SMPL	
CCV48 - 1	CCV48	24828	4.6242	-7470A	PPB	09 Oct 2023 13:13:38	U	SMPL	
CCB48 - 1	CCB48	-60	0.0342	-7470A	PPB	09 Oct 2023 13:15:53	U	SMPL	