



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 10/10/2023

OVENTEMP IN Celsius(°C): 106  
Time IN: 17:00  
In Date: 10/09/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:11  
Out Date: 10/10/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB127692

| Lab ID   | Client SampleID  | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments                 |
|----------|------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|--------------------------|
| O4805-01 | B-1 (0-5)        | 1      | 1.13           | 8.80         | 9.93                    | 8.4                       | 82.6    |                          |
| O4805-02 | O4805-01MS       | 2      | 1.13           | 8.80         | 9.93                    | 8.4                       | 82.6    |                          |
| O4805-03 | O4805-01MSD      | 3      | 1.13           | 8.80         | 9.93                    | 8.4                       | 82.6    |                          |
| O4805-04 | H-1 (0-5)        | 4      | 1.15           | 8.43         | 9.58                    | 8.55                      | 87.8    |                          |
| O4805-05 | M-2 (0-5)        | 5      | 1.15           | 8.47         | 9.62                    | 8.72                      | 89.4    |                          |
| O4805-06 | M-1 (0-5)        | 6      | 1.17           | 8.60         | 9.77                    | 8.85                      | 89.3    |                          |
| O4805-07 | DUP-02           | 7      | 1.2            | 8.52         | 9.72                    | 8.68                      | 87.8    |                          |
| O4812-01 | VNJ-223          | 8      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | STONE SAMPLE,100% SOLIDS |
| O4813-01 | SB-15            | 9      | 1.13           | 8.35         | 9.48                    | 9.42                      | 99.3    |                          |
| O4813-02 | SB-16            | 10     | 1.12           | 8.86         | 9.98                    | 9.94                      | 99.5    |                          |
| O4813-03 | SB-17            | 11     | 1.15           | 8.49         | 9.64                    | 9.45                      | 97.8    |                          |
| O4813-04 | SB-18            | 12     | 1.13           | 8.56         | 9.69                    | 9.64                      | 99.4    |                          |
| O4813-05 | SB-19            | 13     | 1.15           | 8.51         | 9.66                    | 9.63                      | 99.6    |                          |
| O4813-06 | SB-20            | 14     | 1.15           | 8.80         | 9.95                    | 9.86                      | 99.0    |                          |
| O4813-07 | S-9              | 15     | 1.14           | 8.78         | 9.92                    | 9.85                      | 99.2    |                          |
| O4813-08 | S-10             | 16     | 1.19           | 8.41         | 9.6                     | 9.4                       | 97.6    |                          |
| O4813-09 | 50126-50127-COMP | 17     | 1.17           | 8.58         | 9.75                    | 8.95                      | 90.7    |                          |
| O4814-01 | CON-1            | 18     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE SAMPLE          |
| O4816-01 | T-1              | 19     | 1.16           | 8.49         | 9.65                    | 8.65                      | 88.2    |                          |
| O4816-02 | T-2              | 20     | 1.18           | 8.73         | 9.91                    | 9.02                      | 89.8    |                          |
| O4816-03 | T-3              | 21     | 1.12           | 8.46         | 9.58                    | 8.74                      | 90.1    |                          |
| O4816-04 | T-4              | 22     | 1.19           | 8.50         | 9.69                    | 8.84                      | 90.0    |                          |
| O4816-05 | T-5              | 23     | 1.19           | 8.51         | 9.7                     | 8.77                      | 89.1    |                          |
| O4816-06 | T-6              | 24     | 1.15           | 8.81         | 9.96                    | 8.95                      | 88.5    |                          |
| O4816-07 | T-7              | 25     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |                          |
| O4817-01 | 72-11930         | 26     | 1.17           | 8.59         | 9.76                    | 8.62                      | 86.7    |                          |
| O4817-02 | 72-11930-EPH-2   | 27     | 1.19           | 8.38         | 9.57                    | 8.32                      | 85.1    |                          |
| O4817-04 | VNJ-241          | 28     | 1.18           | 8.49         | 9.67                    | 8.66                      | 88.1    |                          |



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 10/10/2023

OVENTEMP IN Celsius(°C): 106  
Time IN: 17:00  
In Date: 10/09/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:11  
Out Date: 10/10/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB127692

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O4817-05 | VNJ-241-EPH-2   | 29     | 1.17           | 8.66         | 9.83                    | 8.79                      | 88.0    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$